

血细胞分析仪用质控物均值及偏差范围

检测机型:LH750

E1166-02

失效期: 2022-01-10

适用机型	测定参数	N均值	偏差范围		
1501L	WBC 10 ⁹ /L	2.0	1.7	—	2.3
	RBC 10 ¹² /L	2.36	2.12	—	2.60
	HGB g/L	61	55	—	67
	HCT %	18.6	15.8	—	21.4
	MCV fL	79.0	71.1	—	86.9
	MCH pg	25.8	20.6	—	31.0
	MCHC g/L	327	278	—	376
	PLT 10 ⁹ /L	77	62	—	92
1501N	WBC 10 ⁹ /L	9.8	8.3	—	11.3
	RBC 10 ¹² /L	4.64	4.18	—	5.10
	HGB g/L	136	122	—	150
	HCT %	40.6	34.5	—	46.7
	MCV fL	87.5	78.8	—	96.3
	MCH pg	29.3	23.4	—	35.2
	MCHC g/L	335	285	—	385
	PLT 10 ⁹ /L	242	194	—	290
1501H	WBC 10 ⁹ /L	18.5	15.7	—	21.3
	RBC 10 ¹² /L	5.80	5.22	—	6.38
	HGB g/L	181	163	—	199
	HCT %	54.2	46.1	—	62.3
	MCV fL	93.5	84.2	—	102.9
	MCH pg	31.2	25.0	—	37.4
	MCHC g/L	334	284	—	384
	PLT 10 ⁹ /L	460	368	—	552

本品定值结果仅供参考。测定值可因仪器型号、试剂等的不同而有改变。各实验室应建立自己的靶值及控制范围。

