

血细胞分析仪用质控物均值及偏差范围

检测机型:LH750

E1162-02

失效期: 2021-07-10

适用机型	测定参数	N均值	偏差范围		
1501L	WBC 10 ⁹ /L	1.9	1.6	—	2.2
	RBC 10 ¹² /L	2.48	2.23	—	2.73
	HGB g/L	61	55	—	67
	HCT %	18.7	15.9	—	21.5
	MCV fL	75.5	68.0	—	83.1
	MCH pg	24.6	19.7	—	29.5
	MCHC g/L	326	277	—	375
	PLT 10 ⁹ /L	69	55	—	83
1501N	WBC 10 ⁹ /L	10.1	8.6	—	11.6
	RBC 10 ¹² /L	4.61	4.15	—	5.07
	HGB g/L	134	121	—	147
	HCT %	40.3	34.3	—	46.3
	MCV fL	87.5	78.8	—	96.3
	MCH pg	29.1	23.3	—	34.9
	MCHC g/L	332	282	—	382
	PLT 10 ⁹ /L	239	191	—	287
1501H	WBC 10 ⁹ /L	20.0	17.0	—	23.0
	RBC 10 ¹² /L	5.53	4.98	—	6.08
	HGB g/L	178	160	—	196
	HCT %	53.4	45.4	—	61.4
	MCV fL	96.5	86.9	—	106.2
	MCH pg	32.2	25.8	—	38.6
	MCHC g/L	334	284	—	384
	PLT 10 ⁹ /L	475	380	—	570

本品定值结果仅供参考。测定值可因仪器型号、试剂等的不同而有改变。各实验室应建立自己的靶值及控制范围。

