

血细胞分析仪用质控物均值及偏差范围

检测机型:LH750		E0021		失效期: 2025-07-16		
适用机型	测定参数	N均值	5.833	偏差范围		
1501L	WBC 10 <sup>9</sup> /L	1.9	1.51	1.6	—	2.2
	RBC 10 <sup>12</sup> /L	2.46	1.90	2.21	—	2.71
	HGB g/L	62	1.00	56	—	68
	HCT %	17.9	1.51	15.2	—	20.6
	MCV fL	72.9	1.00	65.6	—	80.2
	MCH pg	25.2	1.02	20.2	—	30.2
	MCHC g/L	346	1.05	294	—	398
	PLT 10 <sup>9</sup> /L	55	1.02	44	—	66
1501N	WBC 10 <sup>9</sup> /L	9.9	1.51	8.4	—	11.4
	RBC 10 <sup>12</sup> /L	4.52	1.90	4.07	—	4.97
	HGB g/L	135	1.00	122	—	149
	HCT %	40.7	1.51	34.6	—	46.8
	MCV fL	90.0	1.00	81.0	—	99.0
	MCH pg	29.9	1.02	23.9	—	35.9
	MCHC g/L	332	1.05	282	—	382
	PLT 10 <sup>9</sup> /L	249	1.02	199	—	299
1501H	WBC 10 <sup>9</sup> /L	20.4	1.51	17.3	—	23.5
	RBC 10 <sup>12</sup> /L	5.68	1.90	5.11	—	6.25
	HGB g/L	180	1.00	162	—	198
	HCT %	54.8	1.51	46.6	—	63.0
	MCV fL	96.4	1.00	86.8	—	106.0
	MCH pg	31.7	1.02	25.4	—	38.0
	MCHC g/L	329	1.05	280	—	378
	PLT 10 <sup>9</sup> /L	502	1.02	402	—	602

本品定值结果仅供参考。测定值可因仪器型号、试剂等的不同而有改变。各实验室应建立自己的靶值及控制范围。