

血细胞分析仪用质控物均值及偏差范围

检测机型:LH750

E0027

失效期: 2026-01-16

适用机型	测定参数	N均值	偏差范围		
1501L	WBC 10 ⁹ /L	2.0	1.7	—	2.3
	RBC 10 ¹² /L	2.22	2.00	—	2.44
	HGB g/L	60	54	—	66
	HCT %	17.3	14.7	—	19.9
	MCV fL	77.8	70.0	—	85.6
	MCH pg	27.0	21.6	—	32.4
	MCHC g/L	347	295	—	399
	PLT 10 ⁹ /L	56	45	—	67
1501N	WBC 10 ⁹ /L	10.3	8.8	—	11.8
	RBC 10 ¹² /L	4.44	4.00	—	4.88
	HGB g/L	134	121	—	147
	HCT %	39.5	33.6	—	45.4
	MCV fL	88.9	80.0	—	97.8
	MCH pg	30.2	24.2	—	36.2
	MCHC g/L	339	288	—	390
	PLT 10 ⁹ /L	239	191	—	287
1501H	WBC 10 ⁹ /L	20.5	17.4	—	23.6
	RBC 10 ¹² /L	5.69	5.12	—	6.26
	HGB g/L	178	160	—	196
	HCT %	55.3	47.0	—	63.6
	MCV fL	97.2	87.5	—	106.9
	MCH pg	31.3	25.0	—	37.6
	MCHC g/L	322	274	—	370
	PLT 10 ⁹ /L	484	387	—	581

本品定值结果仅供参考。测定值可因仪器型号、试剂等的不同而有改变。各实验室应建立自己的靶值及控制范围。