

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  D1-CRP D3-CRP D5-CRP DH51CRP DH53CRP DH56CRP (适用于技术文件 版本A9.0至A11.9 版本)	WBC	×10 <sup>9</sup> /L	3.47 ±0.50	8.02 ±1.00	18.24 ±2.50
	Neu%	%	49.1 ±9.0	55.6 ±8.0	64.7 ±7.0
	Lym%	%	39.7 ±9.0	29.2 ±8.0	19.0 ±6.0
	Mon%	%	6.3 ±4.0	7.2 ±5.0	6.1 ±6.0
	Eos%	%	4.9 ±4.9	8.0 ±6.0	10.2 ±7.0
	Bas%	%	61.8 ±8.0	72.3 ±8.0	81.6 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.70 ±0.40	4.46 ±0.70	11.80 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.38 ±0.40	2.34 ±0.70	3.47 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.22 ±0.14	0.58 ±0.50	1.11 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.17 ±0.15	0.64 ±0.50	1.86 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.14 ±0.30	5.80 ±0.70	14.88 ±1.50
	RBC	×10 <sup>12</sup> /L	2.24 ±0.18	4.47 ±0.24	5.17 ±0.50
	HGB	g/L	56 ±4	130 ±6	163 ±8
	HCT	%	18.6 ±2.0	41.9 ±3.0	52.6 ±4.0
	MCV	fL	83.0 ±5.0	93.7 ±5.0	101.8 ±6.0
	MCH	pg	25.4 ±2.5	29.4 ±2.5	32.0 ±2.5
	MCHC	g/L	311 ±30	318 ±30	319 ±30
	RDW-CV	%	18.4 ±3.0	16.4 ±3.0	15.7 ±3.0
	RDW-SD	fL	53.5 ±10.0	53.7 ±10.0	55.4 ±12.0
	PLT	×10 <sup>9</sup> /L	52 ±20	270 ±40	553 ±60
	MPV	fL	9.2 ±3.0	9.0 ±3.0	8.3 ±3.0
	PDW	fL	8.8 ±3.0	10.2 ±3.0	8.8 ±3.0
	PCT	%	0.048 ±0.048	0.243 ±0.100	0.459 ±0.200
	P-LCR	%	18.4 ±8.0	20.3 ±8.0	16.0 ±8.0
	P-LCC	×10 <sup>9</sup> /L	11 ±11	55 ±25	89 ±35
	PDW	/	9.9 ±3.0	10.1 ±3.0	10.1 ±3.0

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  D1-CRP D3-CRP D5-CRP DH51CRP DH53CRP DH56CRP (适用于技术文件 版本A11.11及以上 版本)	WBC	×10 <sup>9</sup> /L	3.74 ±0.50	8.83 ±1.00	20.11 ±2.50
	Neu%	%	49.0 ±9.0	55.0 ±8.0	66.1 ±7.0
	Lym%	%	40.0 ±9.0	29.7 ±8.0	19.6 ±6.0
	Mon%	%	6.6 ±4.0	7.5 ±5.0	6.0 ±6.0
	Eos%	%	4.4 ±4.4	7.8 ±6.0	8.3 ±7.0
	Bas%	%	62.5 ±8.0	72.7 ±8.0	82.0 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.83 ±0.40	4.86 ±0.70	13.29 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.50 ±0.40	2.62 ±0.70	3.94 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.25 ±0.14	0.66 ±0.50	1.21 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.16 ±0.15	0.69 ±0.50	1.67 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.34 ±0.30	6.42 ±0.70	16.49 ±1.50
	RBC	×10 <sup>12</sup> /L	2.26 ±0.18	4.50 ±0.24	5.23 ±0.50
	HGB	g/L	55 ±4	128 ±6	162 ±8
	HCT	%	18.1 ±2.0	40.4 ±3.0	50.7 ±4.0
	MCV	fL	80.3 ±5.0	89.8 ±5.0	97.0 ±6.0
	MCH	pg	24.4 ±2.5	28.5 ±2.5	30.9 ±2.5
	MCHC	g/L	310 ±30	323 ±30	325 ±30
	RDW-CV	%	18.1 ±3.0	16.0 ±3.0	15.3 ±3.0
	RDW-SD	fL	52.3 ±10.0	52.2 ±10.0	53.5 ±12.0
	PLT	×10 <sup>9</sup> /L	52 ±20	259 ±40	529 ±60
	MPV	fL	9.2 ±3.0	8.9 ±3.0	8.1 ±3.0
	PDW	fL	8.1 ±3.0	9.9 ±3.0	8.4 ±3.0
	PCT	%	0.048 ±0.048	0.231 ±0.100	0.428 ±0.200
	P-LCR	%	18.0 ±8.0	19.4 ±8.0	15.0 ±8.0
	P-LCC	×10 <sup>9</sup> /L	10 ±10	50 ±25	79 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  UN71 UN73 UN76 (适用于技术文件 版本A5.5至A6.2版 本)	WBC	×10 <sup>9</sup> /L	3.61 ±0.50	8.15 ±1.00	18.47 ±2.50
	Neu%	%	50.0 ±9.0	56.9 ±8.0	65.9 ±7.0
	Lym%	%	40.2 ±9.0	30.1 ±8.0	19.9 ±6.0
	Mon%	%	6.3 ±4.0	7.0 ±5.0	5.9 ±5.9
	Eos%	%	3.5 ±3.5	6.0 ±6.0	8.3 ±7.0
	Bas%	%	0.7 ±0.7	1.0 ±1.0	0.8 ±0.8
	Neu#	×10 <sup>9</sup> /L	1.80 ±0.40	4.64 ±0.70	12.17 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.45 ±0.40	2.45 ±0.70	3.68 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.23 ±0.14	0.57 ±0.50	1.09 ±1.09
	Eos#	×10 <sup>9</sup> /L	0.13 ±0.13	0.49 ±0.49	1.53 ±1.30
	Bas#	×10 <sup>9</sup> /L	0.03 ±0.03	0.08 ±0.08	0.15 ±0.15
	RBC	×10 <sup>12</sup> /L	2.25 ±0.18	4.51 ±0.24	5.21 ±0.50
	HGB	g/L	58 ±4	133 ±6	167 ±8
	HCT	%	18.0 ±2.0	40.7 ±3.0	50.8 ±4.0
	MCV	fL	80.1 ±5.0	90.3 ±5.0	97.6 ±6.0
	MCH	pg	25.5 ±2.5	29.3 ±2.5	32.0 ±2.5
	MCHC	g/L	322 ±30	327 ±30	331 ±30
	RDW-CV	%	18.7 ±3.0	16.8 ±3.0	16.2 ±3.0
	RDW-SD	fL	55.8 ±10.0	56.4 ±10.0	58.5 ±12.0
	PLT	×10 <sup>9</sup> /L	62 ±20	259 ±40	507 ±60
	MPV	fL	8.3 ±3.0	8.9 ±3.0	8.2 ±3.0
	PDW	fL	9.4 ±3.0	9.9 ±3.0	8.5 ±3.0
	PCT	%	0.051 ±0.050	0.231 ±0.100	0.416 ±0.200
	P-LCR	%	16.8 ±8.0	19.9 ±8.0	15.9 ±8.0
	P-LCC	×10 <sup>9</sup> /L	10 ±10	52 ±25	81 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  UN71 UN73 UN76 (适用于技术文件 版本A6.3及以上版本)	WBC	×10 <sup>9</sup> /L	3.70 ±0.50	8.37 ±1.00	19.34 ±2.50
	Neu%	%	50.0 ±9.0	56.8 ±8.0	65.8 ±7.0
	Lym%	%	39.9 ±9.0	29.8 ±8.0	19.8 ±6.0
	Mon%	%	6.4 ±4.0	7.1 ±5.0	5.8 ±5.8
	Eos%	%	3.7 ±3.7	6.3 ±6.0	8.6 ±7.0
	Bas%	%	0.8 ±0.8	0.9 ±0.9	0.9 ±0.9
	Neu#	×10 <sup>9</sup> /L	1.84 ±0.40	4.76 ±0.70	12.73 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.48 ±0.40	2.49 ±0.70	3.83 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.24 ±0.14	0.59 ±0.50	1.12 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.14 ±0.14	0.53 ±0.50	1.66 ±1.30
	Bas#	×10 <sup>9</sup> /L	0.03 ±0.03	0.08 ±0.08	0.17 ±0.17
	RBC	×10 <sup>12</sup> /L	2.29 ±0.18	4.56 ±0.24	5.26 ±0.50
	HGB	g/L	59 ±4	135 ±6	171 ±8
	HCT	%	18.5 ±2.0	41.8 ±3.0	52.1 ±4.0
	MCV	fL	80.7 ±5.0	91.6 ±5.0	99.1 ±6.0
	MCH	pg	25.1 ±2.5	28.9 ±2.5	31.9 ±2.5
	MCHC	g/L	317 ±30	323 ±30	329 ±30
	RDW-CV	%	18.0 ±3.0	16.3 ±3.0	15.6 ±3.0
	RDW-SD	fL	54.1 ±10.0	54.5 ±10.0	56.1 ±12.0
	PLT	×10 <sup>9</sup> /L	57 ±20	254 ±40	519 ±60
	MPV	fL	8.6 ±3.0	9.0 ±3.0	8.1 ±3.0
	PDW	fL	9.9 ±3.0	9.9 ±3.0	8.4 ±3.0
	PCT	%	0.049 ±0.049	0.229 ±0.100	0.420 ±0.200
	P-LCR	%	18.7 ±8.0	20.5 ±8.0	15.4 ±8.0
	P-LCC	×10 <sup>9</sup> /L	11 ±11	51 ±25	81 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DH73 Vet (适用于技术文件 版本B5.5及以上版本)	RBC	×10 <sup>12</sup> /L	2.37 ±0.18	4.74 ±0.24	5.51 ±0.50
	HGB	g/L	58 ±4	133 ±6	169 ±8
	HCT	%	19.5 ±2.0	44.2 ±3.0	55.6 ±4.0
	MCV	fL	82.2 ±5.0	93.2 ±5.0	100.9 ±6.0
	MCH	pg	24.3 ±2.5	28.1 ±2.5	30.8 ±2.5
	MCHC	g/L	300 ±30	306 ±30	309 ±30
	RDW-CV	%	18.0 ±3.0	16.1 ±3.0	15.5 ±3.0
	RDW-SD	fL	51.8 ±10.0	51.9 ±10.0	53.6 ±12.0
	PLT	×10 <sup>9</sup> /L	64 ±20	263 ±40	512 ±60
	MPV	fL	8.6 ±3.0	9.0 ±3.0	8.2 ±3.0
	PDW	fL	9.7 ±3.0	10.2 ±3.0	8.6 ±3.0
	PCT	%	0.055 ±0.050	0.237 ±0.100	0.420 ±0.200
	P-LCR	%	19.0 ±8.0	20.5 ±8.0	16.0 ±8.0
	P-LCC	×10 <sup>9</sup> /L	12 ±12	54 ±25	82 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low		Normal		High	
DYMIND  DF50 DF51 DF52 DF53 DF55 DF56 (适用于技术文件 版本A11.2至A11.6 版本)	WBC	×10 <sup>9</sup> /L	3.56	±0.50	8.07	±1.00	18.31	±2.50
	Neu%	%	48.2	±9.0	54.9	±8.0	63.4	±7.0
	Lym%	%	41.4	±9.0	31.0	±8.0	19.4	±6.0
	Mon%	%	4.6	±4.0	4.8	±4.8	5.7	±5.7
	Eos%	%	5.8	±5.0	9.3	±6.0	11.5	±7.0
	Bas%	%	0.7	±0.7	0.9	±0.9	0.8	±0.8
	Neu#	×10 <sup>9</sup> /L	1.72	±0.40	4.43	±0.70	11.61	±1.40
	Lym#	×10 <sup>9</sup> /L	1.47	±0.40	2.50	±0.70	3.55	±1.10
	Mon#	×10 <sup>9</sup> /L	0.16	±0.14	0.39	±0.39	1.04	±1.04
	Eos#	×10 <sup>9</sup> /L	0.21	±0.15	0.75	±0.50	2.11	±1.30
	Bas#	×10 <sup>9</sup> /L	0.02	±0.02	0.07	±0.07	0.15	±0.15
	RBC	×10 <sup>12</sup> /L	2.29	±0.18	4.55	±0.24	5.22	±0.50
	HGB	g/L	57	±4	132	±6	167	±8
	HCT	%	18.6	±2.0	41.6	±3.0	51.6	±4.0
	MCV	fL	81.1	±5.0	91.5	±5.0	98.9	±6.0
	MCH	pg	24.8	±2.5	28.8	±2.5	31.9	±2.5
	MCHC	g/L	309	±30	319	±30	326	±30
	RDW-CV	%	14.9	±3.0	13.6	±3.0	13.0	±3.0
	RDW-SD	fL	50.6	±10.0	51.2	±10.0	52.8	±12.0
	PLT	×10 <sup>9</sup> /L	68	±20	262	±40	495	±60
	MPV	fL	9.0	±3.0	9.2	±3.0	8.4	±3.0
	PDW	fL	10.2	±3.0	12.0	±3.0	10.5	±3.0
	PCT	%	0.061	±0.050	0.241	±0.100	0.416	±0.200
	P-LCR	%	27.0	±8.0	30.5	±8.0	24.7	±8.0
	P-LCC	×10 <sup>9</sup> /L	18	±15	80	±25	122	±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DF50 DF51 DF52 DF53 DF55 DF56 (适用于技术文件 版本A12.0及以上 版本)	WBC	×10 <sup>9</sup> /L	3.65 ±0.50	8.33 ±1.00	18.91 ±2.50
	Neu%	%	49.3 ±9.0	53.4 ±8.0	64.6 ±7.0
	Lym%	%	40.9 ±9.0	32.5 ±8.0	17.1 ±6.0
	Mon%	%	4.4 ±4.0	5.1 ±5.0	7.3 ±6.0
	Eos%	%	5.4 ±5.0	9.0 ±6.0	11.0 ±7.0
	Bas%	%	0.5 ±0.5	0.5 ±0.5	0.4 ±0.4
	Neu#	×10 <sup>9</sup> /L	1.80 ±0.40	4.45 ±0.70	12.22 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.49 ±0.40	2.71 ±0.70	3.23 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.16 ±0.14	0.42 ±0.42	1.38 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.20 ±0.15	0.75 ±0.50	2.08 ±1.30
	Bas#	×10 <sup>9</sup> /L	0.02 ±0.02	0.04 ±0.04	0.08 ±0.08
	RBC	×10 <sup>12</sup> /L	2.30 ±0.18	4.59 ±0.24	5.28 ±0.50
	HGB	g/L	58 ±4	132 ±6	169 ±8
	HCT	%	18.4 ±2.0	40.6 ±3.0	50.1 ±4.0
	MCV	fL	80.1 ±5.0	88.5 ±5.0	94.9 ±6.0
	MCH	pg	25.0 ±2.5	28.9 ±2.5	32.1 ±2.5
	MCHC	g/L	315 ±30	330 ±30	342 ±30
	RDW-CV	%	15.6 ±3.0	14.1 ±3.0	13.6 ±3.0
	RDW-SD	fL	52.7 ±10.0	53.5 ±10.0	55.4 ±12.0
	PLT	×10 <sup>9</sup> /L	64 ±20	259 ±40	484 ±60
	MPV	fL	8.3 ±3.0	8.7 ±3.0	8.0 ±3.0
	PDW	fL	9.1 ±3.0	11.2 ±3.0	9.7 ±3.0
	PCT	%	0.053 ±0.050	0.225 ±0.100	0.387 ±0.200
	P-LCR	%	21.5 ±8.0	27.1 ±8.0	22.0 ±8.0
	P-LCC	×10 <sup>9</sup> /L	14 ±14	70 ±25	106 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DF50 DF51 DF52 DF53 DF55 DF56 (适用于技术文件 版本B1.0及以上版本)	WBC	×10 <sup>9</sup> /L	3.56 ±0.50	8.14 ±1.00	18.46 ±2.50
	Neu%	%	48.3 ±9.0	55.5 ±8.0	64.1 ±7.0
	Lym%	%	41.4 ±9.0	30.2 ±8.0	20.3 ±6.0
	Mon%	%	4.9 ±4.0	5.3 ±5.0	4.6 ±4.6
	Eos%	%	5.4 ±5.0	9.0 ±6.0	11.0 ±7.0
	Bas%	%	0.6 ±0.6	0.7 ±0.7	0.3 ±0.3
	Neu#	×10 <sup>9</sup> /L	1.73 ±0.40	4.52 ±0.70	11.83 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.47 ±0.40	2.46 ±0.70	3.75 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.17 ±0.14	0.43 ±0.43	0.85 ±0.85
	Eos#	×10 <sup>9</sup> /L	0.19 ±0.15	0.73 ±0.50	2.03 ±1.30
	Bas#	×10 <sup>9</sup> /L	0.02 ±0.02	0.06 ±0.06	0.06 ±0.06
	RBC	×10 <sup>12</sup> /L	2.25 ±0.18	4.52 ±0.24	5.18 ±0.50
	HGB	g/L	58 ±4	135 ±6	171 ±8
	HCT	%	18.1 ±2.0	40.5 ±3.0	49.7 ±4.0
	MCV	fL	80.5 ±5.0	89.5 ±5.0	95.9 ±6.0
	MCH	pg	25.9 ±2.5	30.1 ±2.5	33.4 ±2.5
	MCHC	g/L	325 ±30	340 ±30	352 ±30
	RDW-CV	%	15.2 ±3.0	13.8 ±3.0	13.2 ±3.0
	RDW-SD	fL	51.5 ±10.0	52.1 ±10.0	53.8 ±12.0
	PLT	×10 <sup>9</sup> /L	65 ±20	260 ±40	485 ±60
	MPV	fL	8.2 ±3.0	8.4 ±3.0	7.8 ±3.0
	PDW	fL	9.0 ±3.0	11.4 ±3.0	10.1 ±3.0
	PCT	%	0.053 ±0.050	0.218 ±0.100	0.378 ±0.200
	P-LCR	%	23.1 ±8.0	28.1 ±8.0	22.9 ±8.0
	P-LCC	×10 <sup>9</sup> /L	15 ±15	73 ±25	111 ±35



# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DH71 DH73 DH76 (适用于技术文件 版本A4.1至A10.3 版本)  DH51 DH53 DH56 (适用于技术文件 版本A8.1及以上版 本)	WBC	×10 <sup>9</sup> /L	3.70 ±0.50	8.38 ±1.00	18.59 ±2.50
	Neu%	%	50.9 ±9.0	56.7 ±8.0	66.1 ±7.0
	Lym%	%	39.5 ±9.0	29.1 ±8.0	19.3 ±6.0
	Mon%	%	6.1 ±4.0	7.3 ±5.0	6.1 ±6.0
	Eos%	%	3.5 ±3.5	6.9 ±6.0	8.5 ±7.0
	Bas%	%	61.3 ±8.0	71.8 ±8.0	81.3 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.88 ±0.40	4.75 ±0.70	12.29 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.46 ±0.40	2.44 ±0.70	3.59 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.23 ±0.14	0.61 ±0.50	1.13 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.13 ±0.13	0.58 ±0.50	1.58 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.27 ±0.30	6.02 ±0.70	15.11 ±1.50
	RBC	×10 <sup>12</sup> /L	2.13 ±0.18	4.29 ±0.24	5.02 ±0.50
	HGB	g/L	57 ±4	132 ±6	167 ±8
	HCT	%	17.6 ±2.0	40.0 ±3.0	50.8 ±4.0
	MCV	fL	82.4 ±5.0	93.2 ±5.0	101.1 ±6.0
	MCH	pg	26.0 ±2.5	30.1 ±2.5	32.6 ±2.5
	MCHC	g/L	322 ±30	327 ±30	327 ±30
	RDW-CV	%	18.0 ±3.0	16.0 ±3.0	15.5 ±3.0
	RDW-SD	fL	52.1 ±10.0	52.0 ±10.0	54.1 ±12.0
	PLT	×10 <sup>9</sup> /L	52 ±20	251 ±40	505 ±60
	MPV	fL	9.3 ±3.0	9.2 ±3.0	8.5 ±3.0
	PDW	fL	9.5 ±3.0	10.5 ±3.0	9.1 ±3.0
	PCT	%	0.048 ±0.048	0.231 ±0.100	0.429 ±0.200
	P-LCR	%	20.1 ±8.0	21.3 ±8.0	16.9 ±8.0
	P-LCC	×10 <sup>9</sup> /L	10 ±10	53 ±25	86 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DH71 DH73 DH76 (适用于技术文件 版本 A10.4版本)	WBC	×10 <sup>9</sup> /L	3.43 ±0.50	7.64 ±1.00	16.75 ±2.50
	Neu%	%	49.0 ±9.0	55.8 ±8.0	65.0 ±7.0
	Lym%	%	40.6 ±9.0	30.1 ±8.0	19.8 ±6.0
	Mon%	%	6.5 ±4.0	7.3 ±5.0	6.2 ±6.0
	Eos%	%	3.9 ±3.9	6.8 ±6.0	9.0 ±7.0
	Bas%	%	59.7 ±8.0	70.9 ±8.0	80.7 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.69 ±0.40	4.26 ±0.70	10.88 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.39 ±0.40	2.30 ±0.70	3.32 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.22 ±0.14	0.56 ±0.50	1.04 ±1.04
	Eos#	×10 <sup>9</sup> /L	0.13 ±0.13	0.52 ±0.50	1.51 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.05 ±0.30	5.42 ±0.70	13.52 ±1.50
	RBC	×10 <sup>12</sup> /L	2.23 ±0.18	4.53 ±0.24	5.31 ±0.50
	HGB	g/L	55 ±4	129 ±6	162 ±8
	HCT	%	17.9 ±2.0	40.7 ±3.0	51.5 ±4.0
	MCV	fL	80.1 ±5.0	89.9 ±5.0	97.0 ±6.0
	MCH	pg	22.8 ±2.5	26.3 ±2.5	28.3 ±2.5
	MCHC	g/L	288 ±30	296 ±30	295 ±30
	RDW-CV	%	18.1 ±3.0	16.2 ±3.0	15.5 ±3.0
	RDW-SD	fL	53.0 ±10.0	53.0 ±10.0	54.7 ±12.0
	PLT	×10 <sup>9</sup> /L	55 ±20	255 ±40	514 ±60
	MPV	fL	9.3 ±3.0	9.1 ±3.0	8.3 ±3.0
	PDW	fL	9.5 ±3.0	10.4 ±3.0	9.0 ±3.0
	PCT	%	0.051 ±0.050	0.232 ±0.100	0.427 ±0.200
	P-LCR	%	20.7 ±8.0	20.8 ±8.0	16.6 ±8.0
	P-LCC	×10 <sup>9</sup> /L	11 ±11	53 ±25	85 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DH71 DH73 DH76 (适用于技术文件 版本 A10.5及以上 版本)	WBC	×10 <sup>9</sup> /L	3.81 ±0.50	8.68 ±1.00	19.70 ±2.50
	Neu%	%	49.0 ±9.0	55.7 ±8.0	64.9 ±7.0
	Lym%	%	40.2 ±9.0	29.8 ±8.0	19.6 ±6.0
	Mon%	%	6.6 ±4.0	7.4 ±5.0	6.2 ±6.0
	Eos%	%	4.2 ±4.2	7.1 ±6.0	9.3 ±7.0
	Bas%	%	62.6 ±8.0	72.8 ±8.0	82.4 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.87 ±0.40	4.83 ±0.70	12.79 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.53 ±0.40	2.59 ±0.70	3.86 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.25 ±0.14	0.64 ±0.50	1.22 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.16 ±0.15	0.62 ±0.50	1.83 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.39 ±0.30	6.32 ±0.70	16.23 ±1.50
	RBC	×10 <sup>12</sup> /L	2.28 ±0.18	4.63 ±0.24	5.37 ±0.50
	HGB	g/L	57 ±4	134 ±6	170 ±8
	HCT	%	18.1 ±2.0	41.4 ±3.0	51.9 ±4.0
	MCV	fL	79.5 ±5.0	89.5 ±5.0	96.7 ±6.0
	MCH	pg	24.9 ±2.5	28.7 ±2.5	31.3 ±2.5
	MCHC	g/L	318 ±30	326 ±30	328 ±30
	RDW-CV	%	18.2 ±3.0	16.1 ±3.0	15.4 ±3.0
	RDW-SD	fL	52.6 ±10.0	52.3 ±10.0	53.9 ±12.0
	PLT	×10 <sup>9</sup> /L	51 ±20	262 ±40	517 ±60
	MPV	fL	9.1 ±3.0	8.8 ±3.0	8.0 ±3.0
	PDW	fL	8.5 ±3.0	9.7 ±3.0	8.2 ±3.0
	PCT	%	0.046 ±0.046	0.231 ±0.100	0.414 ±0.200
	P-LCR	%	17.8 ±8.0	19.2 ±8.0	14.6 ±8.0
	P-LCC	×10 <sup>9</sup> /L	11 ±11	50 ±25	75 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  D2-CRP D7-CRP DH71CRP DH73CRP (适用于技术文件 版本A6.0至A6.5版 本)	WBC	×10 <sup>9</sup> /L	3.72 ±0.50	8.48 ±1.00	19.04 ±2.50
	Neu%	%	50.0 ±9.0	55.8 ±8.0	65.1 ±7.0
	Lym%	%	40.4 ±9.0	29.9 ±8.0	19.6 ±6.0
	Mon%	%	6.6 ±4.0	7.3 ±5.0	6.2 ±6.0
	Eos%	%	3.0 ±3.0	7.0 ±6.0	9.1 ±7.0
	Bas%	%	62.4 ±8.0	73.0 ±8.0	82.1 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.86 ±0.40	4.73 ±0.70	12.40 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.50 ±0.40	2.54 ±0.70	3.73 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.25 ±0.14	0.62 ±0.50	1.18 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.11 ±0.11	0.59 ±0.50	1.73 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.32 ±0.30	6.19 ±0.70	15.63 ±1.50
	RBC	×10 <sup>12</sup> /L	2.28 ±0.18	4.61 ±0.24	5.31 ±0.50
	HGB	g/L	57 ±4	134 ±6	168 ±8
	HCT	%	18.4 ±2.0	41.4 ±3.0	51.5 ±4.0
	MCV	fL	80.7 ±5.0	89.9 ±5.0	96.9 ±6.0
	MCH	pg	25.3 ±2.5	29.4 ±2.5	32.0 ±2.5
	MCHC	g/L	318 ±30	331 ±30	333 ±30
	RDW-CV	%	18.4 ±3.0	16.4 ±3.0	15.7 ±3.0
	RDW-SD	fL	53.3 ±10.0	53.6 ±10.0	55.3 ±12.0
	PLT	×10 <sup>9</sup> /L	52 ±20	266 ±40	544 ±60
	MPV	fL	9.3 ±3.0	9.2 ±3.0	8.4 ±3.0
	PDW	fL	8.8 ±3.0	10.6 ±3.0	9.0 ±3.0
	PCT	%	0.048 ±0.048	0.245 ±0.100	0.457 ±0.200
	P-LCR	%	19.3 ±8.0	21.4 ±8.0	16.6 ±8.0
	P-LCC	×10 <sup>9</sup> /L	11 ±11	57 ±25	90 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  D2-CRP D7-CRP DH71CRP DH73CRP (适用于技术文件 版本A6.6及以上版本)	WBC	×10 <sup>9</sup> /L	3.97 ±0.50	8.96 ±1.00	20.41 ±2.50
	Neu%	%	48.8 ±9.0	55.2 ±8.0	64.1 ±7.0
	Lym%	%	40.0 ±9.0	29.6 ±8.0	19.7 ±6.0
	Mon%	%	6.4 ±4.0	7.3 ±5.0	6.1 ±6.0
	Eos%	%	4.8 ±4.8	7.9 ±6.0	10.1 ±7.0
	Bas%	%	62.8 ±8.0	73.1 ±8.0	82.7 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.94 ±0.40	4.95 ±0.70	13.08 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.59 ±0.40	2.65 ±0.70	4.02 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.25 ±0.14	0.65 ±0.50	1.25 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.19 ±0.15	0.71 ±0.50	2.06 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.49 ±0.30	6.55 ±0.70	16.88 ±1.50
	RBC	×10 <sup>12</sup> /L	2.32 ±0.18	4.65 ±0.24	5.34 ±0.50
	HGB	g/L	58 ±4	135 ±6	170 ±8
	HCT	%	18.4 ±2.0	41.3 ±3.0	51.4 ±4.0
	MCV	fL	79.4 ±5.0	88.9 ±5.0	96.3 ±6.0
	MCH	pg	25.1 ±2.5	29.1 ±2.5	31.8 ±2.5
	MCHC	g/L	319 ±30	331 ±30	334 ±30
	RDW-CV	%	18.1 ±3.0	16.1 ±3.0	15.3 ±3.0
	RDW-SD	fL	52.4 ±10.0	52.2 ±10.0	53.6 ±12.0
	PLT	×10 <sup>9</sup> /L	53 ±20	258 ±40	525 ±60
	MPV	fL	8.8 ±3.0	8.6 ±3.0	7.8 ±3.0
	PDW	fL	7.9 ±3.0	9.2 ±3.0	7.7 ±3.0
	PCT	%	0.047 ±0.047	0.222 ±0.100	0.410 ±0.200
	P-LCR	%	16.2 ±8.0	17.6 ±8.0	13.5 ±8.0
	P-LCC	×10 <sup>9</sup> /L	11 ±11	46 ±25	71 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  D6-CRP DH76CRP (适用于技术文件 版本B1.0至B2.0版本)	WBC	×10 <sup>9</sup> /L	3.70 ±0.50	8.33 ±1.00	18.64 ±2.50
	Neu%	%	49.8 ±9.0	56.2 ±8.0	65.6 ±7.0
	Lym%	%	39.9 ±9.0	29.7 ±8.0	19.6 ±6.0
	Mon%	%	6.7 ±4.0	7.7 ±5.0	6.3 ±6.0
	Eos%	%	3.6 ±3.6	6.4 ±6.0	8.5 ±7.0
	Bas%	%	62.2 ±8.0	72.6 ±8.0	82.0 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.84 ±0.40	4.69 ±0.70	12.24 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.48 ±0.40	2.47 ±0.70	3.65 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.25 ±0.14	0.64 ±0.50	1.17 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.13 ±0.13	0.53 ±0.50	1.58 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.30 ±0.30	6.05 ±0.70	15.28 ±1.50
	RBC	×10 <sup>12</sup> /L	2.25 ±0.18	4.59 ±0.24	5.24 ±0.50
	HGB	g/L	59 ±4	134 ±6	168 ±8
	HCT	%	17.9 ±2.0	40.8 ±3.0	50.3 ±4.0
	MCV	fL	79.7 ±5.0	88.9 ±5.0	95.9 ±6.0
	MCH	pg	26.6 ±2.5	29.4 ±2.5	32.2 ±2.5
	MCHC	g/L	337 ±30	333 ±30	339 ±30
	RDW-CV	%	18.3 ±3.0	16.3 ±3.0	15.6 ±3.0
	RDW-SD	fL	52.5 ±10.0	52.6 ±10.0	54.2 ±12.0
	PLT	×10 <sup>9</sup> /L	58 ±20	266 ±40	533 ±60
	MPV	fL	8.1 ±3.0	8.6 ±3.0	7.8 ±3.0
	PDW	fL	8.7 ±3.0	9.4 ±3.0	7.9 ±3.0
	PCT	%	0.047 ±0.047	0.229 ±0.100	0.416 ±0.200
	P-LCR	%	16.0 ±8.0	17.9 ±8.0	13.7 ±8.0
	P-LCC	×10 <sup>9</sup> /L	11 ±11	48 ±25	73 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  D6-CRP DH76CRP (适用于技术文件 版本B2.2及以上版本)	WBC	×10 <sup>9</sup> /L	3.94 ±0.50	8.84 ±1.00	19.77 ±2.50
	Neu%	%	48.9 ±9.0	55.4 ±8.0	66.4 ±7.0
	Lym%	%	39.9 ±9.0	29.9 ±8.0	19.5 ±6.0
	Mon%	%	6.3 ±4.0	7.2 ±5.0	6.0 ±6.0
	Eos%	%	4.9 ±4.9	7.5 ±6.0	8.1 ±7.0
	Bas%	%	62.9 ±8.0	72.9 ±8.0	82.3 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.93 ±0.40	4.90 ±0.70	13.12 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.57 ±0.40	2.64 ±0.70	3.86 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.25 ±0.14	0.64 ±0.50	1.19 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.19 ±0.15	0.66 ±0.50	1.60 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.48 ±0.30	6.44 ±0.70	16.27 ±1.50
	RBC	×10 <sup>12</sup> /L	2.35 ±0.18	4.73 ±0.24	5.41 ±0.50
	HGB	g/L	60 ±4	135 ±6	168 ±8
	HCT	%	18.9 ±2.0	42.6 ±3.0	52.7 ±4.0
	MCV	fL	80.6 ±5.0	90.1 ±5.0	97.4 ±6.0
	MCH	pg	25.6 ±2.5	28.8 ±2.5	31.4 ±2.5
	MCHC	g/L	324 ±30	325 ±30	329 ±30
	RDW-CV	%	18.2 ±3.0	16.2 ±3.0	15.3 ±3.0
	RDW-SD	fL	53.1 ±10.0	53.0 ±10.0	54.4 ±12.0
	PLT	×10 <sup>9</sup> /L	57 ±20	267 ±40	542 ±60
	MPV	fL	8.3 ±3.0	8.8 ±3.0	8.0 ±3.0
	PDW	fL	9.2 ±3.0	9.8 ±3.0	8.4 ±3.0
	PCT	%	0.047 ±0.047	0.235 ±0.100	0.434 ±0.200
	P-LCR	%	17.1 ±8.0	19.4 ±8.0	14.8 ±8.0
	P-LCC	×10 <sup>9</sup> /L	11 ±11	52 ±25	81 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DF50 Vet DF52 Vet DF55 Vet DF56 Vet (适用于技术文件 版本A8.0及以上版 本、B1.0及以上版 本)	WBC	×10 <sup>9</sup> /L	3.58 ±0.50	8.18 ±1.00	18.53 ±2.50
	Neu%	%	48.5 ±9.0	54.0 ±8.0	62.5 ±7.0
	Lym%	%	40.7 ±9.0	31.7 ±8.0	19.5 ±8.0
	Mon%	%	4.8 ±4.0	5.0 ±5.0	6.5 ±6.5
	Eos%	%	6.0 ±5.0	9.3 ±6.0	11.5 ±7.0
	Bas%	%	0.5 ±0.5	0.5 ±0.5	0.5 ±0.5
	Neu#	×10 <sup>9</sup> /L	1.74 ±0.40	4.42 ±0.70	11.59 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.46 ±0.40	2.59 ±0.70	3.61 ±1.40
	Mon#	×10 <sup>9</sup> /L	0.17 ±0.14	0.41 ±0.41	1.20 ±1.20
	Eos#	×10 <sup>9</sup> /L	0.21 ±0.15	0.76 ±0.50	2.13 ±1.30
	Bas#	×10 <sup>9</sup> /L	0.02 ±0.02	0.04 ±0.04	0.09 ±0.09
	RBC	×10 <sup>12</sup> /L	2.34 ±0.18	4.63 ±0.24	5.32 ±0.50
	HGB	g/L	58 ±4	136 ±6	173 ±8
	HCT	%	18.6 ±2.0	41.8 ±3.0	52.2 ±4.0
	MCV	fL	79.7 ±5.0	90.3 ±5.0	98.1 ±6.0
	MCH	pg	24.7 ±2.5	29.4 ±2.5	32.8 ±2.5
	MCHC	g/L	315 ±30	330 ±30	340 ±30
	RDW-CV	%	14.2 ±3.0	13.0 ±3.0	12.4 ±3.0
	RDW-SD	fL	48.0 ±10.0	48.8 ±10.0	50.3 ±12.0
	PLT	×10 <sup>9</sup> /L	66 ±20	258 ±40	480 ±60
	MPV	fL	8.2 ±3.0	8.2 ±3.0	7.7 ±3.0
	PDW	fL	9.1 ±3.0	10.7 ±3.0	9.4 ±3.0
	PCT	%	0.054 ±0.050	0.212 ±0.100	0.370 ±0.200



# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DF1-CRP DF3-CRP DF5-CRP DF50CRP DF52CRP DF53CRP (适用于技术文件 版本A6.1及以上版 本)	WBC	×10 <sup>9</sup> /L	3.86 ±0.50	8.90 ±1.00	20.44 ±2.50
	Neu%	%	48.5 ±9.0	53.9 ±8.0	63.4 ±7.0
	Lym%	%	40.8 ±9.0	31.8 ±8.0	18.4 ±6.0
	Mon%	%	5.2 ±4.0	5.2 ±5.0	7.2 ±6.0
	Eos%	%	5.5 ±5.0	9.1 ±6.0	11.0 ±7.0
	Bas%	%	0.4 ±0.4	0.6 ±0.6	0.4 ±0.4
	Neu#	×10 <sup>9</sup> /L	1.88 ±0.40	4.80 ±0.70	12.96 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.57 ±0.40	2.83 ±0.70	3.76 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.20 ±0.14	0.46 ±0.46	1.47 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.21 ±0.15	0.81 ±0.50	2.25 ±1.30
	Bas#	×10 <sup>9</sup> /L	0.02 ±0.02	0.05 ±0.05	0.08 ±0.08
	RBC	×10 <sup>12</sup> /L	2.22 ±0.18	4.41 ±0.24	5.06 ±0.50
	HGB	g/L	57 ±4	131 ±6	166 ±8
	HCT	%	17.8 ±2.0	39.2 ±3.0	48.3 ±4.0
	MCV	fL	80.3 ±5.0	88.8 ±5.0	95.4 ±6.0
	MCH	pg	25.7 ±2.5	29.5 ±2.5	32.7 ±2.5
	MCHC	g/L	325 ±30	337 ±30	348 ±30
	RDW-CV	%	16.1 ±3.0	14.8 ±3.0	14.2 ±3.0
	RDW-SD	fL	54.7 ±10.0	56.3 ±10.0	58.5 ±12.0
	PLT	×10 <sup>9</sup> /L	49 ±20	261 ±40	540 ±60
	MPV	fL	8.2 ±3.0	8.0 ±3.0	7.5 ±3.0
	PDW	fL	7.9 ±3.0	10.0 ±3.0	8.8 ±3.0
	PCT	%	0.040 ±0.040	0.209 ±0.100	0.405 ±0.200
	P-LCR	%	19.1 ±8.0	22.4 ±8.0	18.6 ±8.0
	P-LCC	×10 <sup>9</sup> /L	11 ±11	59 ±25	100 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low		Normal		High	
DYMIND  DM71X DM72X DM78X DM79X  (适用于技术文件 版本A1.0及以上版本)	WBC	×10 <sup>9</sup> /L	3.73	±0.50	8.39	±1.00	18.80	±2.50
	Neu%	%	50.1	±9.0	55.3	±8.0	64.0	±7.0
	Lym%	%	39.2	±9.0	29.2	±8.0	19.2	±6.0
	Mon%	%	6.5	±4.0	7.3	±5.0	6.0	±6.0
	Eos%	%	4.2	±4.2	8.2	±6.0	10.8	±7.0
	Bas%	%	61.6	±8.0	72.2	±8.0	81.8	±8.0
	Neu#	×10 <sup>9</sup> /L	1.87	±0.40	4.64	±0.70	12.03	±1.40
	Lym#	×10 <sup>9</sup> /L	1.46	±0.40	2.45	±0.70	3.61	±1.10
	Mon#	×10 <sup>9</sup> /L	0.24	±0.14	0.61	±0.50	1.13	±1.10
	Eos#	×10 <sup>9</sup> /L	0.16	±0.15	0.69	±0.50	2.03	±1.30
	Bas#	×10 <sup>9</sup> /L	2.30	±0.30	6.06	±0.70	15.38	±1.50
	RBC	×10 <sup>12</sup> /L	2.27	±0.18	4.59	±0.24	5.29	±0.50
	HGB	g/L	60	±4	134	±6	168	±8
	HCT	%	17.9	±2.0	41.0	±3.0	51.3	±4.0
	MCV	fL	78.7	±5.0	89.3	±5.0	96.9	±6.0
	MCH	pg	26.2	±2.5	29.2	±2.5	31.7	±2.5
	MCHC	g/L	337	±30	330	±30	330	±30
	RDW-CV	%	18.4	±3.0	16.4	±3.0	15.6	±3.0
	RDW-SD	fL	52.7	±10.0	52.9	±10.0	54.6	±12.0
	PLT	×10 <sup>9</sup> /L	54	±20	260	±40	541	±60
	MPV	fL	8.1	±3.0	8.6	±3.0	7.9	±3.0
	PDW	fL	8.9	±3.0	9.5	±3.0	8.1	±3.0
	PCT	%	0.044	±0.044	0.224	±0.100	0.427	±0.200
	P-LCR	%	15.4	±8.0	17.9	±8.0	13.9	±8.0
	P-LCC	×10 <sup>9</sup> /L	10	±10	46	±25	75	±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  UN71 Vet UN73 Vet (适用于技术文件 版本A1.0及以上版本)	WBC	×10 <sup>9</sup> /L	3.67 ±0.50	8.29 ±1.00	19.07 ±2.50
	Neu%	%	49.6 ±9.0	57.2 ±8.0	66.6 ±7.0
	Lym%	%	39.9 ±9.0	30.2 ±8.0	20.3 ±6.0
	Mon%	%	6.1 ±4.0	7.2 ±5.0	5.9 ±5.9
	Eos%	%	4.4 ±4.4	5.4 ±5.4	7.2 ±7.0
	Bas%	%	0.7 ±0.7	0.9 ±0.9	0.9 ±0.9
	Neu#	×10 <sup>9</sup> /L	1.83 ±0.40	4.74 ±0.70	12.70 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.46 ±0.40	2.50 ±0.70	3.87 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.22 ±0.14	0.60 ±0.50	1.13 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.16 ±0.15	0.45 ±0.45	1.37 ±1.30
	Bas#	×10 <sup>9</sup> /L	0.03 ±0.03	0.07 ±0.07	0.17 ±0.17
	RBC	×10 <sup>12</sup> /L	2.31 ±0.18	4.55 ±0.24	5.22 ±0.50
	HGB	g/L	60 ±4	133 ±6	167 ±8
	HCT	%	18.2 ±2.0	40.8 ±3.0	50.7 ±4.0
	MCV	fL	79.0 ±5.0	89.6 ±5.0	97.2 ±6.0
	MCH	pg	25.7 ±2.5	29.2 ±2.5	31.9 ±2.5
	MCHC	g/L	329 ±30	329 ±30	332 ±30
	RDW-CV	%	18.5 ±3.0	16.4 ±3.0	15.5 ±3.0
	RDW-SD	fL	46.3 ±10.0	46.1 ±10.0	47.2 ±12.0
	PLT	×10 <sup>9</sup> /L	48 ±20	248 ±40	481 ±60
	MPV	fL	8.8 ±3.0	8.1 ±3.0	7.6 ±3.0
	PDW	fL	7.1 ±3.0	8.7 ±3.0	7.3 ±3.0
	PCT	%	0.042 ±0.042	0.201 ±0.100	0.366 ±0.200
	P-LCR	%	14.5 ±8.0	15.9 ±8.0	12.9 ±8.0
	P-LCC	×10 <sup>9</sup> /L	10 ±10	39 ±25	62 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5213-02

EXP. 2025-05-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DM60 Vet DM61 Vet DM62 Vet DM63 Vet (适用于技术文件 版本A1.4及以上版本)	WBC	×10 <sup>9</sup> /L	3.27 ±0.50	7.40 ±1.00	16.95 ±2.50
	Neu%	%	48.9 ±9.0	54.8 ±8.0	63.5 ±7.0
	Lym%	%	40.8 ±9.0	30.4 ±8.0	20.1 ±6.0
	Mon%	%	4.7 ±4.0	5.5 ±5.0	5.0 ±5.0
	Eos%	%	5.6 ±5.0	9.3 ±6.0	11.4 ±7.0
	Bas%	%	0.5 ±0.5	0.5 ±0.5	0.4 ±0.4
	Neu#	×10 <sup>9</sup> /L	1.61 ±0.40	4.05 ±0.70	10.76 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.33 ±0.40	2.25 ±0.70	3.41 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.15 ±0.14	0.41 ±0.41	0.85 ±0.85
	Eos#	×10 <sup>9</sup> /L	0.18 ±0.15	0.69 ±0.50	1.93 ±1.30
	Bas#	×10 <sup>9</sup> /L	0.02 ±0.02	0.04 ±0.04	0.07 ±0.07
	RBC	×10 <sup>12</sup> /L	2.16 ±0.18	4.27 ±0.24	4.94 ±0.50
	HGB	g/L	55 ±4	130 ±6	167 ±8
	HCT	%	17.2 ±2.0	37.8 ±3.0	46.9 ±4.0
	MCV	fL	79.4 ±5.0	88.5 ±5.0	95.0 ±6.0
	MCH	pg	25.3 ±2.5	30.0 ±2.5	33.4 ±2.5
	MCHC	g/L	323 ±30	343 ±30	356 ±30
	RDW-CV	%	14.2 ±3.0	12.8 ±3.0	12.2 ±3.0
	RDW-SD	fL	50.6 ±10.0	51.2 ±10.0	52.7 ±12.0
	PLT	×10 <sup>9</sup> /L	57 ±20	244 ±40	458 ±60
	MPV	fL	8.2 ±3.0	8.2 ±3.0	7.6 ±3.0
	PDW	fL	8.2 ±3.0	10.5 ±3.0	9.1 ±3.0
	PCT	%	0.047 ±0.047	0.200 ±0.100	0.348 ±0.200