

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  D1-CRP D3-CRP D5-CRP DH51CRP DH53CRP DH56CRP (适用于技术文件 版本A9.0至A11.9 版本)	WBC	×10 <sup>9</sup> /L	3.47 ±0.50	8.15 ±1.00	19.14 ±2.50
	Neu%	%	51.1 ±9.0	58.3 ±8.0	64.6 ±7.0
	Lym%	%	36.3 ±9.0	28.9 ±8.0	20.2 ±6.0
	Mon%	%	7.6 ±4.0	6.6 ±5.0	6.8 ±6.0
	Eos%	%	5.0 ±5.0	6.2 ±6.0	8.4 ±7.0
	Bas%	%	64.8 ±8.0	71.5 ±8.0	80.5 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.78 ±0.40	4.74 ±0.70	12.36 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.26 ±0.40	2.36 ±0.70	3.87 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.26 ±0.14	0.54 ±0.50	1.30 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.17 ±0.15	0.51 ±0.50	1.61 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.25 ±0.30	5.83 ±0.70	15.41 ±1.50
	RBC	×10 <sup>12</sup> /L	2.25 ±0.18	4.61 ±0.24	5.35 ±0.50
	HGB	g/L	59 ±4	132 ±6	167 ±8
	HCT	%	19.6 ±2.0	42.2 ±3.0	52.8 ±4.0
	MCV	fL	86.9 ±5.0	91.6 ±5.0	98.7 ±6.0
	MCH	pg	25.9 ±2.5	28.4 ±2.5	30.9 ±2.5
	MCHC	g/L	304 ±30	315 ±30	319 ±30
	RDW-CV	%	18.0 ±3.0	16.5 ±3.0	16.1 ±3.0
	RDW-SD	fL	54.3 ±10.0	52.5 ±10.0	54.6 ±12.0
	PLT	×10 <sup>9</sup> /L	53 ±20	259 ±40	554 ±60
	MPV	fL	9.0 ±3.0	8.4 ±3.0	8.2 ±3.0
	PDW	fL	7.6 ±3.0	8.9 ±3.0	8.4 ±3.0
	PCT	%	0.048 ±0.048	0.218 ±0.100	0.454 ±0.200
	P-LCR	%	17.0 ±8.0	17.5 ±8.0	16.0 ±8.0
	P-LCC	×10 <sup>9</sup> /L	11 ±11	45 ±25	88 ±35
	PDW	/	10.1 ±3.0	10.1 ±3.0	10.1 ±3.0

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  D1-CRP D3-CRP D5-CRP DH51CRP DH53CRP DH56CRP (适用于技术文件 版本A11.11及以上 版本)	WBC	×10 <sup>9</sup> /L	3.79 ±0.50	8.90 ±1.00	20.93 ±2.50
	Neu%	%	50.7 ±9.0	57.2 ±8.0	64.4 ±7.0
	Lym%	%	36.6 ±9.0	29.9 ±8.0	20.5 ±6.0
	Mon%	%	7.8 ±4.0	7.1 ±5.0	7.0 ±6.0
	Eos%	%	4.9 ±4.9	5.8 ±5.8	8.1 ±7.0
	Bas%	%	65.3 ±8.0	72.0 ±8.0	80.7 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.91 ±0.40	5.09 ±0.70	13.47 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.39 ±0.40	2.66 ±0.70	4.29 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.30 ±0.14	0.63 ±0.50	1.47 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.19 ±0.15	0.52 ±0.50	1.70 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.47 ±0.30	6.41 ±0.70	16.89 ±1.50
	RBC	×10 <sup>12</sup> /L	2.25 ±0.18	4.65 ±0.24	5.36 ±0.50
	HGB	g/L	57 ±4	129 ±6	164 ±8
	HCT	%	19.0 ±2.0	41.0 ±3.0	50.8 ±4.0
	MCV	fL	84.5 ±5.0	88.2 ±5.0	94.7 ±6.0
	MCH	pg	25.2 ±2.5	27.7 ±2.5	30.4 ±2.5
	MCHC	g/L	303 ±30	318 ±30	326 ±30
	RDW-CV	%	17.6 ±3.0	16.1 ±3.0	15.6 ±3.0
	RDW-SD	fL	53.4 ±10.0	51.6 ±10.0	53.4 ±12.0
	PLT	×10 <sup>9</sup> /L	54 ±20	248 ±40	534 ±60
	MPV	fL	9.0 ±3.0	8.3 ±3.0	8.1 ±3.0
	PDW	fL	7.5 ±3.0	8.4 ±3.0	8.0 ±3.0
	PCT	%	0.049 ±0.049	0.206 ±0.100	0.433 ±0.200
	P-LCR	%	16.9 ±8.0	17.1 ±8.0	15.3 ±8.0
	P-LCC	×10 <sup>9</sup> /L	10 ±10	43 ±25	82 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  UN71 UN73 UN76 (适用于技术文件 版本A5.5至A6.2版 本)	WBC	×10 <sup>9</sup> /L	3.54 ±0.50	8.13 ±1.00	18.91 ±2.50
	Neu%	%	51.1 ±9.0	57.3 ±8.0	65.5 ±7.0
	Lym%	%	38.0 ±9.0	31.0 ±8.0	21.2 ±6.0
	Mon%	%	7.1 ±4.0	6.9 ±5.0	6.7 ±6.0
	Eos%	%	3.8 ±3.8	4.8 ±4.8	6.6 ±6.6
	Bas%	%	0.8 ±0.8	1.0 ±1.0	0.8 ±0.8
	Neu#	×10 <sup>9</sup> /L	1.81 ±0.40	4.66 ±0.70	12.38 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.35 ±0.40	2.52 ±0.70	4.01 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.25 ±0.14	0.56 ±0.50	1.27 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.13 ±0.13	0.39 ±0.39	1.25 ±1.25
	Bas#	×10 <sup>9</sup> /L	0.03 ±0.03	0.08 ±0.08	0.15 ±0.15
	RBC	×10 <sup>12</sup> /L	2.23 ±0.18	4.62 ±0.24	5.28 ±0.50
	HGB	g/L	59 ±4	133 ±6	167 ±8
	HCT	%	18.7 ±2.0	40.9 ±3.0	50.2 ±4.0
	MCV	fL	84.0 ±5.0	88.5 ±5.0	95.0 ±6.0
	MCH	pg	25.4 ±2.5	27.6 ±2.5	30.3 ±2.5
	MCHC	g/L	307 ±30	315 ±30	323 ±30
	RDW-CV	%	18.3 ±3.0	17.0 ±3.0	16.5 ±3.0
	RDW-SD	fL	56.8 ±10.0	55.3 ±10.0	57.4 ±12.0
	PLT	×10 <sup>9</sup> /L	63 ±20	244 ±40	500 ±60
	MPV	fL	8.2 ±3.0	8.3 ±3.0	8.2 ±3.0
	PDW	fL	8.8 ±3.0	8.7 ±3.0	8.4 ±3.0
	PCT	%	0.052 ±0.050	0.203 ±0.100	0.410 ±0.200
	P-LCR	%	16.5 ±8.0	17.1 ±8.0	16.0 ±8.0
	P-LCC	×10 <sup>9</sup> /L	10 ±10	42 ±25	80 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  UN71 UN73 UN76 (适用于技术文件 版本A6.3及以上版本)	WBC	×10 <sup>9</sup> /L	3.83 ±0.50	8.59 ±1.00	20.57 ±2.50
	Neu%	%	51.7 ±9.0	57.9 ±8.0	65.7 ±7.0
	Lym%	%	38.0 ±9.0	30.7 ±8.0	21.4 ±6.0
	Mon%	%	6.4 ±4.0	6.5 ±5.0	6.5 ±6.0
	Eos%	%	3.9 ±3.9	4.9 ±4.9	6.4 ±6.4
	Bas%	%	0.6 ±0.6	0.6 ±0.6	0.8 ±0.8
	Neu#	×10 <sup>9</sup> /L	1.97 ±0.40	4.97 ±0.70	13.51 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.46 ±0.40	2.64 ±0.70	4.40 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.25 ±0.14	0.56 ±0.50	1.34 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.15 ±0.15	0.42 ±0.42	1.32 ±1.30
	Bas#	×10 <sup>9</sup> /L	0.02 ±0.02	0.05 ±0.05	0.16 ±0.16
	RBC	×10 <sup>12</sup> /L	2.33 ±0.18	4.78 ±0.24	5.53 ±0.50
	HGB	g/L	62 ±4	138 ±6	174 ±8
	HCT	%	19.8 ±2.0	43.0 ±3.0	53.4 ±4.0
	MCV	fL	84.9 ±5.0	89.9 ±5.0	96.6 ±6.0
	MCH	pg	26.1 ±2.5	28.7 ±2.5	31.5 ±2.5
	MCHC	g/L	312 ±30	323 ±30	330 ±30
	RDW-CV	%	17.5 ±3.0	16.3 ±3.0	15.9 ±3.0
	RDW-SD	fL	55.3 ±10.0	53.6 ±10.0	55.8 ±12.0
	PLT	×10 <sup>9</sup> /L	61 ±20	246 ±40	498 ±60
	MPV	fL	8.3 ±3.0	8.4 ±3.0	8.1 ±3.0
	PDW	fL	8.4 ±3.0	8.5 ±3.0	8.1 ±3.0
	PCT	%	0.051 ±0.050	0.207 ±0.100	0.403 ±0.200
	P-LCR	%	17.5 ±8.0	17.6 ±8.0	15.8 ±8.0
	P-LCC	×10 <sup>9</sup> /L	11 ±11	43 ±25	79 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DH73 Vet (适用于技术文件 版本B5.5及以上版本)	RBC	×10 <sup>12</sup> /L	2.33 ±0.18	4.82 ±0.24	5.60 ±0.50
	HGB	g/L	60 ±4	135 ±6	173 ±8
	HCT	%	20.2 ±2.0	44.4 ±3.0	55.4 ±4.0
	MCV	fL	86.8 ±5.0	92.1 ±5.0	98.9 ±6.0
	MCH	pg	24.8 ±2.5	27.2 ±2.5	29.9 ±2.5
	MCHC	g/L	292 ±30	301 ±30	309 ±30
	RDW-CV	%	17.5 ±3.0	16.2 ±3.0	15.7 ±3.0
	RDW-SD	fL	52.6 ±10.0	50.8 ±10.0	52.6 ±12.0
	PLT	×10 <sup>9</sup> /L	78 ±20	255 ±40	515 ±60
	MPV	fL	7.9 ±3.0	8.4 ±3.0	8.2 ±3.0
	PDW	fL	7.3 ±3.0	8.7 ±3.0	8.3 ±3.0
	PCT	%	0.062 ±0.050	0.214 ±0.100	0.422 ±0.200
	P-LCR	%	15.4 ±8.0	17.7 ±8.0	16.2 ±8.0
	P-LCC	×10 <sup>9</sup> /L	12 ±12	45 ±25	83 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DF50 DF51 DF52 DF53 DF55 DF56 (适用于技术文件 版本A11.2至A11.6 版本)	WBC	×10 <sup>9</sup> /L	3.59 ±0.50	8.23 ±1.00	19.49 ±2.50
	Neu%	%	52.4 ±9.0	57.9 ±8.0	65.1 ±7.0
	Lym%	%	36.5 ±9.0	29.6 ±8.0	19.2 ±6.0
	Mon%	%	5.4 ±4.0	5.1 ±5.0	5.7 ±5.7
	Eos%	%	5.7 ±5.0	7.4 ±6.0	10.0 ±7.0
	Bas%	%	3.4 ±3.4	3.1 ±3.1	5.0 ±5.0
	Neu#	×10 <sup>9</sup> /L	1.89 ±0.40	4.76 ±0.70	12.69 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.31 ±0.40	2.44 ±0.70	3.74 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.19 ±0.14	0.42 ±0.42	1.11 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.20 ±0.15	0.61 ±0.50	1.95 ±1.30
	Bas#	×10 <sup>9</sup> /L	0.12 ±0.12	0.26 ±0.26	0.97 ±0.97
	RBC	×10 <sup>12</sup> /L	2.26 ±0.18	4.65 ±0.24	5.39 ±0.50
	HGB	g/L	59 ±4	133 ±6	172 ±8
	HCT	%	19.2 ±2.0	41.8 ±3.0	52.0 ±4.0
	MCV	fL	85.0 ±5.0	89.9 ±5.0	96.4 ±6.0
	MCH	pg	25.7 ±2.5	28.4 ±2.5	31.7 ±2.5
	MCHC	g/L	308 ±30	322 ±30	335 ±30
	RDW-CV	%	14.6 ±3.0	13.7 ±3.0	13.4 ±3.0
	RDW-SD	fL	51.2 ±10.0	50.2 ±10.0	52.3 ±12.0
	PLT	×10 <sup>9</sup> /L	68 ±20	248 ±40	487 ±60
	MPV	fL	8.7 ±3.0	8.5 ±3.0	8.3 ±3.0
	PDW	fL	9.1 ±3.0	10.5 ±3.0	10.2 ±3.0
	PCT	%	0.059 ±0.050	0.211 ±0.100	0.404 ±0.200
	P-LCR	%	23.6 ±8.0	25.9 ±8.0	24.3 ±8.0
	P-LCC	×10 <sup>9</sup> /L	16 ±15	64 ±25	118 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DF50 DF51 DF52 DF53 DF55 DF56 (适用于技术文件 版本A12.0及以上 版本)	WBC	×10 <sup>9</sup> /L	3.77 ±0.50	8.63 ±1.00	20.44 ±2.50
	Neu%	%	51.5 ±9.0	56.6 ±8.0	64.1 ±7.0
	Lym%	%	38.1 ±9.0	31.4 ±8.0	20.6 ±6.0
	Mon%	%	5.0 ±4.0	4.8 ±4.8	5.7 ±5.7
	Eos%	%	5.4 ±5.0	7.2 ±6.0	9.6 ±7.0
	Bas%	%	2.0 ±2.0	1.5 ±1.5	2.5 ±2.5
	Neu#	×10 <sup>9</sup> /L	1.94 ±0.40	4.89 ±0.70	13.10 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.44 ±0.40	2.71 ±0.70	4.21 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.19 ±0.14	0.41 ±0.41	1.17 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.20 ±0.15	0.62 ±0.50	1.96 ±1.30
	Bas#	×10 <sup>9</sup> /L	0.08 ±0.08	0.13 ±0.13	0.51 ±0.51
	RBC	×10 <sup>12</sup> /L	2.29 ±0.18	4.72 ±0.24	5.46 ±0.50
	HGB	g/L	60 ±4	135 ±6	173 ±8
	HCT	%	19.2 ±2.0	40.9 ±3.0	50.4 ±4.0
	MCV	fL	83.8 ±5.0	86.6 ±5.0	92.3 ±6.0
	MCH	pg	25.8 ±2.5	28.2 ±2.5	31.3 ±2.5
	MCHC	g/L	313 ±30	331 ±30	346 ±30
	RDW-CV	%	15.1 ±3.0	14.3 ±3.0	14.0 ±3.0
	RDW-SD	fL	53.2 ±10.0	52.7 ±10.0	54.9 ±12.0
	PLT	×10 <sup>9</sup> /L	65 ±20	242 ±40	474 ±60
	MPV	fL	8.1 ±3.0	8.0 ±3.0	7.9 ±3.0
	PDW	fL	8.3 ±3.0	9.7 ±3.0	9.5 ±3.0
	PCT	%	0.053 ±0.050	0.194 ±0.100	0.374 ±0.200
	P-LCR	%	20.2 ±8.0	22.8 ±8.0	21.9 ±8.0
	P-LCC	×10 <sup>9</sup> /L	13 ±13	55 ±25	104 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DF50 DF51 DF52 DF53 DF55 DF56 (适用于技术文件 版本B1.0及以上版 本)	WBC	×10 <sup>9</sup> /L	3.64 ±0.50	8.42 ±1.00	19.79 ±2.50
	Neu%	%	53.1 ±9.0	58.7 ±8.0	64.9 ±7.0
	Lym%	%	35.5 ±9.0	28.8 ±8.0	19.7 ±6.0
	Mon%	%	6.2 ±4.0	5.7 ±5.0	5.6 ±5.6
	Eos%	%	5.2 ±5.0	6.8 ±6.0	9.8 ±7.0
	Bas%	%	2.6 ±2.6	2.1 ±2.1	2.6 ±2.6
	Neu#	×10 <sup>9</sup> /L	1.93 ±0.40	4.95 ±0.70	12.84 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.29 ±0.40	2.42 ±0.70	3.90 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.23 ±0.14	0.48 ±0.48	1.11 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.19 ±0.15	0.57 ±0.50	1.94 ±1.30
	Bas#	×10 <sup>9</sup> /L	0.09 ±0.09	0.18 ±0.18	0.51 ±0.51
	RBC	×10 <sup>12</sup> /L	2.26 ±0.18	4.66 ±0.24	5.38 ±0.50
	HGB	g/L	60 ±4	136 ±6	176 ±8
	HCT	%	19.1 ±2.0	40.8 ±3.0	50.2 ±4.0
	MCV	fL	84.4 ±5.0	87.5 ±5.0	93.3 ±6.0
	MCH	pg	26.5 ±2.5	29.4 ±2.5	32.8 ±2.5
	MCHC	g/L	319 ±30	341 ±30	357 ±30
	RDW-CV	%	14.9 ±3.0	13.9 ±3.0	13.7 ±3.0
	RDW-SD	fL	52.2 ±10.0	51.2 ±10.0	53.6 ±12.0
	PLT	×10 <sup>9</sup> /L	66 ±20	249 ±40	486 ±60
	MPV	fL	7.9 ±3.0	7.8 ±3.0	7.7 ±3.0
	PDW	fL	8.4 ±3.0	10.2 ±3.0	9.8 ±3.0
	PCT	%	0.052 ±0.050	0.194 ±0.100	0.374 ±0.200
	P-LCR	%	18.5 ±8.0	21.5 ±8.0	20.6 ±8.0
	P-LCC	×10 <sup>9</sup> /L	12 ±12	54 ±25	100 ±35



# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DH71 DH73 DH76 (适用于技术文件 版本A4.1至A10.3 版本)  DH51 DH53 DH56 (适用于技术文件 版本A8.1及以上版 本)	WBC	×10 <sup>9</sup> /L	3.75 ±0.50	8.53 ±1.00	19.55 ±2.50
	Neu%	%	52.4 ±9.0	57.9 ±8.0	65.6 ±7.0
	Lym%	%	36.4 ±9.0	30.6 ±8.0	20.7 ±6.0
	Mon%	%	6.6 ±4.0	6.5 ±5.0	6.8 ±6.0
	Eos%	%	4.6 ±4.6	5.0 ±5.0	6.9 ±6.9
	Bas%	%	64.6 ±8.0	71.1 ±8.0	79.8 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.96 ±0.40	4.94 ±0.70	12.82 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.37 ±0.40	2.61 ±0.70	4.05 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.25 ±0.14	0.55 ±0.50	1.33 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.17 ±0.15	0.43 ±0.43	1.35 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.42 ±0.30	6.06 ±0.70	15.60 ±1.50
	RBC	×10 <sup>12</sup> /L	2.12 ±0.18	4.40 ±0.24	5.11 ±0.50
	HGB	g/L	59 ±4	135 ±6	171 ±8
	HCT	%	18.3 ±2.0	40.3 ±3.0	50.4 ±4.0
	MCV	fL	86.4 ±5.0	91.6 ±5.0	98.6 ±6.0
	MCH	pg	28.4 ±2.5	31.2 ±2.5	34.1 ±2.5
	MCHC	g/L	333 ±30	343 ±30	349 ±30
	RDW-CV	%	17.5 ±3.0	16.1 ±3.0	15.6 ±3.0
	RDW-SD	fL	52.5 ±10.0	50.8 ±10.0	52.6 ±12.0
	PLT	×10 <sup>9</sup> /L	52 ±20	241 ±40	513 ±60
	MPV	fL	9.3 ±3.0	8.8 ±3.0	8.4 ±3.0
	PDW	fL	8.4 ±3.0	9.1 ±3.0	8.7 ±3.0
	PCT	%	0.048 ±0.048	0.212 ±0.100	0.431 ±0.200
	P-LCR	%	18.9 ±8.0	19.3 ±8.0	17.1 ±8.0
	P-LCC	×10 <sup>9</sup> /L	10 ±10	46 ±25	87 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DH71 DH73 DH76 (适用于技术文件 版本 A10.4版本)	WBC	×10 <sup>9</sup> /L	3.36 ±0.50	7.66 ±1.00	17.48 ±2.50
	Neu%	%	51.1 ±9.0	57.2 ±8.0	64.9 ±7.0
	Lym%	%	37.4 ±9.0	30.9 ±8.0	21.6 ±6.0
	Mon%	%	7.3 ±4.0	6.9 ±5.0	6.8 ±6.0
	Eos%	%	4.2 ±4.2	5.0 ±5.0	6.7 ±6.7
	Bas%	%	63.3 ±8.0	70.0 ±8.0	79.4 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.71 ±0.40	4.38 ±0.70	11.34 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.26 ±0.40	2.37 ±0.70	3.78 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.25 ±0.14	0.53 ±0.50	1.19 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.14 ±0.14	0.38 ±0.38	1.17 ±1.17
	Bas#	×10 <sup>9</sup> /L	2.13 ±0.30	5.36 ±0.70	13.88 ±1.50
	RBC	×10 <sup>12</sup> /L	2.20 ±0.18	4.63 ±0.24	5.40 ±0.50
	HGB	g/L	57 ±4	131 ±6	167 ±8
	HCT	%	18.3 ±2.0	40.5 ±3.0	50.6 ±4.0
	MCV	fL	83.2 ±5.0	87.4 ±5.0	93.7 ±6.0
	MCH	pg	27.1 ±2.5	29.7 ±2.5	32.4 ±2.5
	MCHC	g/L	327 ±30	341 ±30	347 ±30
	RDW-CV	%	17.7 ±3.0	16.2 ±3.0	15.8 ±3.0
	RDW-SD	fL	53.8 ±10.0	51.8 ±10.0	53.8 ±12.0
	PLT	×10 <sup>9</sup> /L	51 ±20	245 ±40	523 ±60
	MPV	fL	9.2 ±3.0	8.6 ±3.0	8.3 ±3.0
	PDW	fL	7.8 ±3.0	9.1 ±3.0	8.7 ±3.0
	PCT	%	0.047 ±0.047	0.211 ±0.100	0.434 ±0.200
	P-LCR	%	18.5 ±8.0	18.6 ±8.0	16.4 ±8.0
	P-LCC	×10 <sup>9</sup> /L	11 ±11	46 ±25	86 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DH71 DH73 DH76 (适用于技术文件 版本 A10.5及以上 版本)	WBC	×10 <sup>9</sup> /L	3.80 ±0.50	8.76 ±1.00	20.53 ±2.50
	Neu%	%	51.0 ±9.0	57.0 ±8.0	64.8 ±7.0
	Lym%	%	37.3 ±9.0	31.1 ±8.0	21.2 ±6.0
	Mon%	%	7.4 ±4.0	6.9 ±5.0	6.9 ±6.0
	Eos%	%	4.3 ±4.3	5.0 ±5.0	7.1 ±7.0
	Bas%	%	65.3 ±8.0	72.4 ±8.0	81.4 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.94 ±0.40	5.00 ±0.70	13.30 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.42 ±0.40	2.72 ±0.70	4.35 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.28 ±0.14	0.60 ±0.50	1.42 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.16 ±0.15	0.44 ±0.44	1.46 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.48 ±0.30	6.34 ±0.70	16.71 ±1.50
	RBC	×10 <sup>12</sup> /L	2.31 ±0.18	4.82 ±0.24	5.59 ±0.50
	HGB	g/L	60 ±4	137 ±6	173 ±8
	HCT	%	19.2 ±2.0	42.3 ±3.0	52.7 ±4.0
	MCV	fL	83.3 ±5.0	87.7 ±5.0	94.2 ±6.0
	MCH	pg	25.9 ±2.5	28.2 ±2.5	30.7 ±2.5
	MCHC	g/L	316 ±30	326 ±30	331 ±30
	RDW-CV	%	17.7 ±3.0	16.4 ±3.0	16.0 ±3.0
	RDW-SD	fL	53.3 ±10.0	51.7 ±10.0	53.8 ±12.0
	PLT	×10 <sup>9</sup> /L	53 ±20	247 ±40	517 ±60
	MPV	fL	8.9 ±3.0	8.3 ±3.0	8.0 ±3.0
	PDW	fL	7.4 ±3.0	8.4 ±3.0	8.0 ±3.0
	PCT	%	0.047 ±0.047	0.205 ±0.100	0.414 ±0.200
	P-LCR	%	17.2 ±8.0	17.0 ±8.0	15.0 ±8.0
	P-LCC	×10 <sup>9</sup> /L	11 ±11	42 ±25	78 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  D2-CRP D7-CRP DH71CRP DH73CRP (适用于技术文件 版本A6.0至A6.5版 本)	WBC	×10 <sup>9</sup> /L	3.72 ±0.50	8.63 ±1.00	19.99 ±2.50
	Neu%	%	51.2 ±9.0	57.0 ±8.0	64.9 ±7.0
	Lym%	%	37.2 ±9.0	31.0 ±8.0	21.2 ±6.0
	Mon%	%	7.5 ±4.0	6.7 ±5.0	6.8 ±6.0
	Eos%	%	4.1 ±4.1	5.3 ±5.3	7.1 ±7.0
	Bas%	%	64.9 ±8.0	72.1 ±8.0	81.0 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.91 ±0.40	4.91 ±0.70	12.97 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.38 ±0.40	2.68 ±0.70	4.24 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.28 ±0.14	0.58 ±0.50	1.36 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.15 ±0.15	0.46 ±0.46	1.42 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.41 ±0.30	6.22 ±0.70	16.19 ±1.50
	RBC	×10 <sup>12</sup> /L	2.27 ±0.18	4.79 ±0.24	5.54 ±0.50
	HGB	g/L	60 ±4	138 ±6	173 ±8
	HCT	%	19.2 ±2.0	42.2 ±3.0	52.3 ±4.0
	MCV	fL	84.8 ±5.0	88.1 ±5.0	94.4 ±6.0
	MCH	pg	25.9 ±2.5	28.2 ±2.5	30.6 ±2.5
	MCHC	g/L	310 ±30	324 ±30	329 ±30
	RDW-CV	%	17.9 ±3.0	16.5 ±3.0	16.1 ±3.0
	RDW-SD	fL	54.2 ±10.0	52.6 ±10.0	54.7 ±12.0
	PLT	×10 <sup>9</sup> /L	56 ±20	252 ±40	551 ±60
	MPV	fL	9.1 ±3.0	8.6 ±3.0	8.2 ±3.0
	PDW	fL	7.9 ±3.0	9.1 ±3.0	8.6 ±3.0
	PCT	%	0.051 ±0.050	0.217 ±0.100	0.452 ±0.200
	P-LCR	%	18.2 ±8.0	18.5 ±8.0	16.1 ±8.0
	P-LCC	×10 <sup>9</sup> /L	10 ±10	47 ±25	88 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  D2-CRP D7-CRP DH71CRP DH73CRP (适用于技术文件 版本A6.6及以上版本)	WBC	×10 <sup>9</sup> /L	3.97 ±0.50	9.08 ±1.00	21.02 ±2.50
	Neu%	%	50.3 ±9.0	56.8 ±8.0	64.2 ±7.0
	Lym%	%	37.5 ±9.0	30.9 ±8.0	21.0 ±6.0
	Mon%	%	7.1 ±4.0	6.7 ±5.0	6.7 ±6.0
	Eos%	%	5.1 ±5.0	5.6 ±5.6	8.1 ±7.0
	Bas%	%	65.6 ±8.0	72.4 ±8.0	81.5 ±8.0
	Neu#	×10 <sup>9</sup> /L	2.00 ±0.40	5.15 ±0.70	13.50 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.49 ±0.40	2.81 ±0.70	4.41 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.28 ±0.14	0.61 ±0.50	1.41 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.20 ±0.15	0.51 ±0.50	1.70 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.60 ±0.30	6.57 ±0.70	17.13 ±1.50
	RBC	×10 <sup>12</sup> /L	2.32 ±0.18	4.82 ±0.24	5.54 ±0.50
	HGB	g/L	61 ±4	138 ±6	174 ±8
	HCT	%	19.4 ±2.0	42.2 ±3.0	52.1 ±4.0
	MCV	fL	83.6 ±5.0	87.5 ±5.0	94.1 ±6.0
	MCH	pg	26.0 ±2.5	28.3 ±2.5	31.1 ±2.5
	MCHC	g/L	315 ±30	328 ±30	336 ±30
	RDW-CV	%	17.5 ±3.0	16.1 ±3.0	15.6 ±3.0
	RDW-SD	fL	52.9 ±10.0	51.1 ±10.0	52.9 ±12.0
	PLT	×10 <sup>9</sup> /L	55 ±20	240 ±40	524 ±60
	MPV	fL	8.7 ±3.0	8.0 ±3.0	7.8 ±3.0
	PDW	fL	6.7 ±3.0	7.8 ±3.0	7.4 ±3.0
	PCT	%	0.048 ±0.048	0.192 ±0.100	0.409 ±0.200
	P-LCR	%	15.2 ±8.0	15.2 ±8.0	13.8 ±8.0
	P-LCC	×10 <sup>9</sup> /L	10 ±10	36 ±25	72 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  D6-CRP DH76CRP (适用于技术文件 版本B1.0至B2.0版本)	WBC	×10 <sup>9</sup> /L	3.62 ±0.50	8.40 ±1.00	19.31 ±2.50
	Neu%	%	50.9 ±9.0	57.0 ±8.0	65.2 ±7.0
	Lym%	%	37.3 ±9.0	30.7 ±8.0	20.9 ±6.0
	Mon%	%	7.7 ±4.0	7.2 ±5.0	7.2 ±6.0
	Eos%	%	4.1 ±4.1	5.1 ±5.1	6.7 ±6.7
	Bas%	%	65.4 ±8.0	71.9 ±8.0	80.9 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.84 ±0.40	4.79 ±0.70	12.59 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.35 ±0.40	2.58 ±0.70	4.04 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.28 ±0.14	0.60 ±0.50	1.39 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.15 ±0.15	0.43 ±0.43	1.29 ±1.29
	Bas#	×10 <sup>9</sup> /L	2.37 ±0.30	6.04 ±0.70	15.62 ±1.50
	RBC	×10 <sup>12</sup> /L	2.25 ±0.18	4.71 ±0.24	5.46 ±0.50
	HGB	g/L	62 ±4	137 ±6	172 ±8
	HCT	%	18.8 ±2.0	41.0 ±3.0	50.9 ±4.0
	MCV	fL	83.5 ±5.0	87.0 ±5.0	93.3 ±6.0
	MCH	pg	27.1 ±2.5	28.5 ±2.5	30.9 ±2.5
	MCHC	g/L	329 ±30	333 ±30	336 ±30
	RDW-CV	%	17.9 ±3.0	16.4 ±3.0	15.9 ±3.0
	RDW-SD	fL	53.2 ±10.0	51.4 ±10.0	53.3 ±12.0
	PLT	×10 <sup>9</sup> /L	60 ±20	247 ±40	534 ±60
	MPV	fL	7.8 ±3.0	8.0 ±3.0	7.8 ±3.0
	PDW	fL	7.9 ±3.0	8.0 ±3.0	7.7 ±3.0
	PCT	%	0.047 ±0.047	0.198 ±0.100	0.417 ±0.200
	P-LCR	%	14.5 ±8.0	15.2 ±8.0	14.0 ±8.0
	P-LCC	×10 <sup>9</sup> /L	11 ±11	38 ±25	75 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  D6-CRP DH76CRP (适用于技术文件 版本B2.2及以上版本)	WBC	×10 <sup>9</sup> /L	3.91 ±0.50	8.93 ±1.00	20.41 ±2.50
	Neu%	%	50.6 ±9.0	56.9 ±8.0	64.4 ±7.0
	Lym%	%	36.6 ±9.0	30.3 ±8.0	20.6 ±6.0
	Mon%	%	7.5 ±4.0	6.9 ±5.0	6.7 ±6.0
	Eos%	%	5.3 ±5.0	5.9 ±5.9	8.3 ±7.0
	Bas%	%	65.1 ±8.0	72.0 ±8.0	81.1 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.98 ±0.40	5.07 ±0.70	13.15 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.43 ±0.40	2.71 ±0.70	4.20 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.29 ±0.14	0.62 ±0.50	1.37 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.21 ±0.15	0.53 ±0.50	1.69 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.55 ±0.30	6.43 ±0.70	16.55 ±1.50
	RBC	×10 <sup>12</sup> /L	2.36 ±0.18	4.90 ±0.24	5.62 ±0.50
	HGB	g/L	61 ±4	135 ±6	169 ±8
	HCT	%	20.0 ±2.0	43.3 ±3.0	53.3 ±4.0
	MCV	fL	84.8 ±5.0	88.4 ±5.0	94.9 ±6.0
	MCH	pg	26.1 ±2.5	27.7 ±2.5	30.2 ±2.5
	MCHC	g/L	312 ±30	317 ±30	323 ±30
	RDW-CV	%	17.6 ±3.0	16.2 ±3.0	15.6 ±3.0
	RDW-SD	fL	54.1 ±10.0	52.3 ±10.0	53.9 ±12.0
	PLT	×10 <sup>9</sup> /L	58 ±20	257 ±40	543 ±60
	MPV	fL	8.2 ±3.0	8.3 ±3.0	8.0 ±3.0
	PDW	fL	8.6 ±3.0	8.3 ±3.0	8.0 ±3.0
	PCT	%	0.048 ±0.048	0.213 ±0.100	0.434 ±0.200
	P-LCR	%	16.9 ±8.0	16.8 ±8.0	15.0 ±8.0
	P-LCC	×10 <sup>9</sup> /L	10 ±10	43 ±25	81 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DF50 Vet DF52 Vet DF55 Vet DF56 Vet (适用于技术文件 版本A8.0及以上版 本、B1.0及以上版 本)	WBC	×10 <sup>9</sup> /L	3.68 ±0.50	8.41 ±1.00	19.84 ±2.50
	Neu%	%	52.0 ±9.0	56.6 ±8.0	65.2 ±7.0
	Lym%	%	37.6 ±9.0	31.7 ±8.0	17.3 ±9.0
	Mon%	%	4.9 ±4.0	4.1 ±4.1	7.6 ±7.0
	Eos%	%	5.5 ±5.0	7.6 ±6.0	9.9 ±8.0
	Bas%	%	2.0 ±2.0	1.4 ±1.4	2.6 ±2.6
	Neu#	×10 <sup>9</sup> /L	1.92 ±0.40	4.76 ±0.70	12.94 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.38 ±0.40	2.67 ±0.70	3.43 ±2.00
	Mon#	×10 <sup>9</sup> /L	0.18 ±0.14	0.34 ±0.34	1.51 ±1.50
	Eos#	×10 <sup>9</sup> /L	0.20 ±0.15	0.64 ±0.50	1.96 ±1.60
	Bas#	×10 <sup>9</sup> /L	0.07 ±0.07	0.12 ±0.12	0.52 ±0.52
	RBC	×10 <sup>12</sup> /L	2.31 ±0.18	4.70 ±0.24	5.47 ±0.50
	HGB	g/L	59 ±4	136 ±6	177 ±8
	HCT	%	19.3 ±2.0	41.7 ±3.0	52.3 ±4.0
	MCV	fL	83.6 ±5.0	88.7 ±5.0	95.6 ±6.0
	MCH	pg	25.5 ±2.5	29.0 ±2.5	32.3 ±2.5
	MCHC	g/L	311 ±30	334 ±30	347 ±30
	RDW-CV	%	13.9 ±3.0	13.1 ±3.0	12.8 ±3.0
	RDW-SD	fL	48.6 ±10.0	47.9 ±10.0	49.8 ±12.0
	PLT	×10 <sup>9</sup> /L	64 ±20	237 ±40	461 ±60
	MPV	fL	7.8 ±3.0	7.7 ±3.0	7.6 ±3.0
	PDW	fL	8.1 ±3.0	9.5 ±3.0	9.3 ±3.0
	PCT	%	0.050 ±0.050	0.182 ±0.100	0.350 ±0.200



# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DF1-CRP DF3-CRP DF5-CRP DF50CRP DF52CRP DF53CRP (适用于技术文件 版本A6.1及以上版 本)	WBC	×10 <sup>9</sup> /L	3.91 ±0.50	9.16 ±1.00	21.86 ±2.50
	Neu%	%	51.8 ±9.0	57.2 ±8.0	64.1 ±7.0
	Lym%	%	36.7 ±9.0	30.6 ±8.0	20.2 ±6.0
	Mon%	%	6.2 ±4.0	5.1 ±5.0	6.2 ±6.0
	Eos%	%	5.3 ±5.0	7.1 ±6.0	9.5 ±7.0
	Bas%	%	2.0 ±2.0	1.7 ±1.7	2.7 ±2.7
	Neu#	×10 <sup>9</sup> /L	2.03 ±0.40	5.24 ±0.70	14.00 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.43 ±0.40	2.80 ±0.70	4.42 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.24 ±0.14	0.47 ±0.47	1.36 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.21 ±0.15	0.65 ±0.50	2.08 ±1.30
	Bas#	×10 <sup>9</sup> /L	0.08 ±0.08	0.16 ±0.16	0.59 ±0.59
	RBC	×10 <sup>12</sup> /L	2.25 ±0.18	4.57 ±0.24	5.25 ±0.50
	HGB	g/L	60 ±4	135 ±6	172 ±8
	HCT	%	18.9 ±2.0	39.6 ±3.0	48.6 ±4.0
	MCV	fL	84.1 ±5.0	86.7 ±5.0	92.5 ±6.0
	MCH	pg	26.7 ±2.5	29.4 ±2.5	32.5 ±2.5
	MCHC	g/L	323 ±30	345 ±30	360 ±30
	RDW-CV	%	16.0 ±3.0	15.6 ±3.0	15.5 ±3.0
	RDW-SD	fL	56.5 ±10.0	57.3 ±10.0	60.9 ±12.0
	PLT	×10 <sup>9</sup> /L	52 ±20	257 ±40	550 ±60
	MPV	fL	7.9 ±3.0	7.4 ±3.0	7.4 ±3.0
	PDW	fL	7.1 ±3.0	8.8 ±3.0	8.5 ±3.0
	PCT	%	0.041 ±0.041	0.190 ±0.100	0.407 ±0.200
	P-LCR	%	17.3 ±8.0	19.1 ±8.0	18.6 ±8.0
	P-LCC	×10 <sup>9</sup> /L	11 ±11	49 ±25	102 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DM71X DM72X DM78X DM79X  (适用于技术文件 版本A1.0及以上版本)	WBC	×10 <sup>9</sup> /L	3.72 ±0.50	8.58 ±1.00	19.66 ±2.50
	Neu%	%	51.0 ±9.0	57.2 ±8.0	64.1 ±7.0
	Lym%	%	36.6 ±9.0	29.9 ±8.0	20.4 ±6.0
	Mon%	%	7.3 ±4.0	7.1 ±5.0	7.0 ±6.0
	Eos%	%	5.1 ±5.0	5.8 ±5.8	8.5 ±7.0
	Bas%	%	64.7 ±8.0	71.5 ±8.0	80.7 ±8.0
	Neu#	×10 <sup>9</sup> /L	1.90 ±0.40	4.90 ±0.70	12.60 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.36 ±0.40	2.57 ±0.70	4.01 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.27 ±0.14	0.61 ±0.50	1.38 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.19 ±0.15	0.50 ±0.50	1.67 ±1.30
	Bas#	×10 <sup>9</sup> /L	2.41 ±0.30	6.13 ±0.70	15.87 ±1.50
	RBC	×10 <sup>12</sup> /L	2.29 ±0.18	4.77 ±0.24	5.53 ±0.50
	HGB	g/L	62 ±4	136 ±6	171 ±8
	HCT	%	19.0 ±2.0	41.9 ±3.0	52.3 ±4.0
	MCV	fL	82.8 ±5.0	87.9 ±5.0	94.6 ±6.0
	MCH	pg	27.2 ±2.5	28.8 ±2.5	31.2 ±2.5
	MCHC	g/L	332 ±30	332 ±30	334 ±30
	RDW-CV	%	17.9 ±3.0	16.5 ±3.0	16.0 ±3.0
	RDW-SD	fL	53.5 ±10.0	52.0 ±10.0	53.9 ±12.0
	PLT	×10 <sup>9</sup> /L	55 ±20	250 ±40	539 ±60
	MPV	fL	7.8 ±3.0	8.0 ±3.0	7.8 ±3.0
	PDW	fL	8.2 ±3.0	8.1 ±3.0	7.6 ±3.0
	PCT	%	0.043 ±0.043	0.200 ±0.100	0.420 ±0.200
	P-LCR	%	14.6 ±8.0	15.5 ±8.0	14.0 ±8.0
	P-LCC	×10 <sup>9</sup> /L	10 ±10	39 ±25	75 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  UN71 Vet UN73 Vet (适用于技术文件 版本A1.0及以上版本)	WBC	×10 <sup>9</sup> /L	3.62 ±0.50	8.31 ±1.00	19.77 ±2.50
	Neu%	%	51.2 ±9.0	57.7 ±8.0	66.2 ±7.0
	Lym%	%	36.9 ±9.0	30.3 ±8.0	20.9 ±6.0
	Mon%	%	8.0 ±4.0	7.0 ±5.0	7.0 ±6.0
	Eos%	%	3.9 ±3.9	5.0 ±5.0	5.9 ±5.9
	Bas%	%	0.8 ±0.8	0.7 ±0.7	0.8 ±0.8
	Neu#	×10 <sup>9</sup> /L	1.85 ±0.40	4.79 ±0.70	13.09 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.34 ±0.40	2.52 ±0.70	4.13 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.29 ±0.14	0.58 ±0.50	1.38 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.14 ±0.14	0.42 ±0.42	1.17 ±1.17
	Bas#	×10 <sup>9</sup> /L	0.03 ±0.03	0.06 ±0.06	0.16 ±0.16
	RBC	×10 <sup>12</sup> /L	2.31 ±0.18	4.70 ±0.24	5.42 ±0.50
	HGB	g/L	62 ±4	135 ±6	171 ±8
	HCT	%	19.2 ±2.0	41.3 ±3.0	51.3 ±4.0
	MCV	fL	83.2 ±5.0	87.9 ±5.0	94.7 ±6.0
	MCH	pg	26.5 ±2.5	28.3 ±2.5	31.1 ±2.5
	MCHC	g/L	323 ±30	327 ±30	334 ±30
	RDW-CV	%	18.0 ±3.0	16.6 ±3.0	16.0 ±3.0
	RDW-SD	fL	47.1 ±10.0	45.4 ±10.0	46.8 ±12.0
	PLT	×10 <sup>9</sup> /L	47 ±20	233 ±40	476 ±60
	MPV	fL	8.7 ±3.0	7.7 ±3.0	7.6 ±3.0
	PDW	fL	6.5 ±3.0	7.4 ±3.0	6.9 ±3.0
	PCT	%	0.041 ±0.041	0.179 ±0.100	0.362 ±0.200
	P-LCR	%	14.1 ±8.0	14.3 ±8.0	13.2 ±8.0
	P-LCC	×10 <sup>9</sup> /L	9 ±9	33 ±25	63 ±35

# HC5D-B

(Cat No.1510)

For MINDRAY  
HEMATOLOGY CONTROLS

CONTROL

LOT E5216

EXP. 2025-07-16

适用机型	参数	单位	Low	Normal	High
DYMIND  DM60 Vet DM61 Vet DM62 Vet DM63 Vet (适用于技术文件 版本A1.4及以上版本)	WBC	×10 <sup>9</sup> /L	3.37 ±0.50	7.79 ±1.00	18.37 ±2.50
	Neu%	%	52.2 ±9.0	57.6 ±8.0	64.0 ±7.0
	Lym%	%	35.7 ±9.0	29.4 ±8.0	19.9 ±6.0
	Mon%	%	6.7 ±4.0	5.9 ±5.0	6.3 ±6.0
	Eos%	%	5.4 ±5.0	7.1 ±6.0	9.8 ±7.0
	Bas%	%	1.8 ±1.8	1.6 ±1.6	2.8 ±2.8
	Neu#	×10 <sup>9</sup> /L	1.76 ±0.40	4.49 ±0.70	11.75 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.20 ±0.40	2.29 ±0.70	3.66 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.23 ±0.14	0.46 ±0.46	1.16 ±1.10
	Eos#	×10 <sup>9</sup> /L	0.18 ±0.15	0.55 ±0.50	1.80 ±1.30
	Bas#	×10 <sup>9</sup> /L	0.06 ±0.06	0.12 ±0.12	0.51 ±0.51
	RBC	×10 <sup>12</sup> /L	2.16 ±0.18	4.41 ±0.24	5.10 ±0.50
	HGB	g/L	58 ±4	133 ±6	171 ±8
	HCT	%	18.1 ±2.0	38.6 ±3.0	47.6 ±4.0
	MCV	fL	83.9 ±5.0	87.5 ±5.0	93.4 ±6.0
	MCH	pg	26.1 ±2.5	29.5 ±2.5	33.0 ±2.5
	MCHC	g/L	325 ±30	346 ±30	363 ±30
	RDW-CV	%	13.8 ±3.0	13.0 ±3.0	12.7 ±3.0
	RDW-SD	fL	51.1 ±10.0	50.5 ±10.0	52.8 ±12.0
	PLT	×10 <sup>9</sup> /L	58 ±20	230 ±40	454 ±60
	MPV	fL	8.0 ±3.0	7.7 ±3.0	7.5 ±3.0
	PDW	fL	7.6 ±3.0	9.3 ±3.0	8.9 ±3.0
	PCT	%	0.046 ±0.046	0.177 ±0.100	0.341 ±0.200