

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND D1-CRP D3-CRP D5-CRP DH51CRP DH53CRP DH56CRP (适用于技术文件 版本A9.0至A11.9 版本)	WBC	×10 ⁹ /L	3.41 ±0.50	8.33 ±1.00	19.30 ±2.50
	Neu%	%	51.4 ±9.0	64.4 ±8.0	64.1 ±7.0
	Lym%	%	35.2 ±9.0	26.0 ±8.0	19.9 ±6.0
	Mon%	%	9.1 ±4.0	3.4 ±3.4	7.6 ±6.0
	Eos%	%	4.3 ±4.3	6.2 ±6.0	8.4 ±7.0
	Bas%	%	63.9 ±8.0	70.6 ±8.0	80.5 ±8.0
	Neu#	×10 ⁹ /L	1.75 ±0.40	5.36 ±0.70	12.37 ±1.40
	Lym#	×10 ⁹ /L	1.20 ±0.40	2.17 ±0.70	3.84 ±1.10
	Mon#	×10 ⁹ /L	0.31 ±0.14	0.28 ±0.28	1.47 ±1.10
	Eos#	×10 ⁹ /L	0.15 ±0.15	0.52 ±0.50	1.62 ±1.30
	Bas#	×10 ⁹ /L	2.18 ±0.30	5.88 ±0.70	15.54 ±1.50
	RBC	×10 ¹² /L	2.30 ±0.18	4.54 ±0.24	5.27 ±0.50
	HGB	g/L	59 ±4	133 ±6	165 ±8
	HCT	%	19.3 ±2.0	41.9 ±3.0	52.1 ±4.0
	MCV	fL	83.9 ±5.0	92.4 ±5.0	98.9 ±6.0
	MCH	pg	25.6 ±2.5	29.2 ±2.5	31.2 ±2.5
	MCHC	g/L	309 ±30	320 ±30	321 ±30
	RDW-CV	%	19.1 ±3.0	16.9 ±3.0	16.4 ±3.0
	RDW-SD	fL	56.4 ±10.0	54.6 ±10.0	56.3 ±12.0
	PLT	×10 ⁹ /L	48 ±20	255 ±40	544 ±60
	MPV	fL	8.8 ±3.0	8.3 ±3.0	8.3 ±3.0
	PDW	fL	7.1 ±3.0	9.0 ±3.0	9.1 ±3.0
	PCT	%	0.042 ±0.042	0.212 ±0.100	0.452 ±0.200
	P-LCR	%	15.9 ±8.0	16.9 ±8.0	16.5 ±8.0
	P-LCC	×10 ⁹ /L	13 ±13	43 ±25	90 ±35
	PDW	/	10.1 ±3.0	9.9 ±3.0	9.9 ±3.0

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND D1-CRP D3-CRP D5-CRP DH51CRP DH53CRP DH56CRP (适用于技术文件 版本A11.11及以上 版本)	WBC	×10 ⁹ /L	3.73 ±0.50	9.23 ±1.00	21.39 ±2.50
	Neu%	%	50.6 ±9.0	55.9 ±8.0	63.3 ±7.0
	Lym%	%	35.3 ±9.0	30.0 ±8.0	20.1 ±6.0
	Mon%	%	9.5 ±4.0	7.9 ±5.0	8.8 ±6.0
	Eos%	%	4.6 ±4.6	6.2 ±6.0	7.8 ±7.0
	Bas%	%	65.2 ±8.0	70.9 ±8.0	80.8 ±8.0
	Neu#	×10 ⁹ /L	1.89 ±0.40	5.16 ±0.70	13.54 ±1.40
	Lym#	×10 ⁹ /L	1.32 ±0.40	2.77 ±0.70	4.30 ±1.10
	Mon#	×10 ⁹ /L	0.35 ±0.14	0.73 ±0.50	1.88 ±1.10
	Eos#	×10 ⁹ /L	0.17 ±0.15	0.57 ±0.50	1.67 ±1.30
	Bas#	×10 ⁹ /L	2.43 ±0.30	6.54 ±0.70	17.28 ±1.50
	RBC	×10 ¹² /L	2.32 ±0.18	4.65 ±0.24	5.37 ±0.50
	HGB	g/L	59 ±4	134 ±6	169 ±8
	HCT	%	19.0 ±2.0	41.6 ±3.0	51.1 ±4.0
	MCV	fL	82.1 ±5.0	89.5 ±5.0	95.1 ±6.0
	MCH	pg	25.2 ±2.5	28.8 ±2.5	30.7 ±2.5
	MCHC	g/L	310 ±30	325 ±30	326 ±30
	RDW-CV	%	18.8 ±3.0	16.5 ±3.0	16.1 ±3.0
	RDW-SD	fL	55.8 ±10.0	53.6 ±10.0	55.2 ±12.0
	PLT	×10 ⁹ /L	52 ±20	244 ±40	523 ±60
	MPV	fL	9.0 ±3.0	8.3 ±3.0	8.2 ±3.0
	PDW	fL	7.5 ±3.0	8.7 ±3.0	8.7 ±3.0
	PCT	%	0.047 ±0.047	0.203 ±0.100	0.429 ±0.200
	P-LCR	%	16.9 ±8.0	16.8 ±8.0	15.9 ±8.0
	P-LCC	×10 ⁹ /L	14 ±14	41 ±25	82 ±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND UN71 UN73 UN76 (适用于技术文件 版本A5.5至A6.2版 本)	WBC	×10 ⁹ /L	3.49 ±0.50	8.27 ±1.00	19.18 ±2.50
	Neu%	%	52.2 ±9.0	56.6 ±8.0	64.6 ±7.0
	Lym%	%	36.1 ±9.0	31.5 ±8.0	21.2 ±6.0
	Mon%	%	8.5 ±4.0	7.3 ±5.0	8.0 ±6.0
	Eos%	%	3.2 ±3.2	4.6 ±4.6	6.2 ±6.2
	Bas%	%	0.7 ±0.7	0.7 ±0.7	0.9 ±0.9
	Neu#	×10 ⁹ /L	1.82 ±0.40	4.68 ±0.70	12.39 ±1.40
	Lym#	×10 ⁹ /L	1.26 ±0.40	2.61 ±0.70	4.07 ±1.10
	Mon#	×10 ⁹ /L	0.30 ±0.14	0.60 ±0.50	1.53 ±1.10
	Eos#	×10 ⁹ /L	0.11 ±0.11	0.38 ±0.38	1.19 ±1.19
	Bas#	×10 ⁹ /L	0.02 ±0.02	0.06 ±0.06	0.17 ±0.17
	RBC	×10 ¹² /L	2.29 ±0.18	4.50 ±0.24	5.22 ±0.50
	HGB	g/L	60 ±4	133 ±6	165 ±8
	HCT	%	18.7 ±2.0	40.2 ±3.0	49.6 ±4.0
	MCV	fL	81.7 ±5.0	89.4 ±5.0	95.1 ±6.0
	MCH	pg	26.4 ±2.5	29.7 ±2.5	31.9 ±2.5
	MCHC	g/L	326 ±30	333 ±30	336 ±30
	RDW-CV	%	19.4 ±3.0	17.2 ±3.0	16.9 ±3.0
	RDW-SD	fL	59.1 ±10.0	57.1 ±10.0	59.2 ±12.0
	PLT	×10 ⁹ /L	59 ±20	239 ±40	493 ±60
	MPV	fL	7.9 ±3.0	8.3 ±3.0	8.3 ±3.0
	PDW	fL	8.6 ±3.0	8.8 ±3.0	8.9 ±3.0
	PCT	%	0.047 ±0.047	0.198 ±0.100	0.409 ±0.200
	P-LCR	%	15.2 ±8.0	16.9 ±8.0	16.6 ±8.0
	P-LCC	×10 ⁹ /L	14 ±14	40 ±25	82 ±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND UN71 UN73 UN76 (适用于技术文件 版本A6.3及以上版本)	WBC	×10 ⁹ /L	3.64 ±0.50	8.82 ±1.00	20.55 ±2.50
	Neu%	%	51.5 ±9.0	57.0 ±8.0	64.4 ±7.0
	Lym%	%	36.4 ±9.0	32.1 ±8.0	21.8 ±6.0
	Mon%	%	7.6 ±4.0	6.2 ±5.0	7.5 ±6.0
	Eos%	%	4.5 ±4.5	4.7 ±4.7	6.3 ±6.3
	Bas%	%	0.8 ±0.8	0.7 ±0.7	0.8 ±0.8
	Neu#	×10 ⁹ /L	1.88 ±0.40	5.03 ±0.70	13.24 ±1.40
	Lym#	×10 ⁹ /L	1.32 ±0.40	2.83 ±0.70	4.48 ±1.10
	Mon#	×10 ⁹ /L	0.28 ±0.14	0.55 ±0.50	1.54 ±1.10
	Eos#	×10 ⁹ /L	0.16 ±0.15	0.41 ±0.41	1.29 ±1.29
	Bas#	×10 ⁹ /L	0.03 ±0.03	0.06 ±0.06	0.16 ±0.16
	RBC	×10 ¹² /L	2.38 ±0.18	4.68 ±0.24	5.40 ±0.50
	HGB	g/L	61 ±4	135 ±6	167 ±8
	HCT	%	19.6 ±2.0	42.6 ±3.0	52.3 ±4.0
	MCV	fL	82.2 ±5.0	91.0 ±5.0	96.8 ±6.0
	MCH	pg	25.6 ±2.5	28.8 ±2.5	30.9 ±2.5
	MCHC	g/L	311 ±30	317 ±30	319 ±30
	RDW-CV	%	18.6 ±3.0	16.4 ±3.0	16.1 ±3.0
	RDW-SD	fL	57.6 ±10.0	55.1 ±10.0	57.3 ±12.0
	PLT	×10 ⁹ /L	57 ±20	242 ±40	491 ±60
	MPV	fL	8.1 ±3.0	8.4 ±3.0	8.3 ±3.0
	PDW	fL	8.8 ±3.0	8.9 ±3.0	8.8 ±3.0
	PCT	%	0.046 ±0.046	0.203 ±0.100	0.408 ±0.200
	P-LCR	%	16.7 ±8.0	17.2 ±8.0	16.1 ±8.0
	P-LCC	×10 ⁹ /L	15 ±15	42 ±25	79 ±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND DH73 Vet (适用于技术文件 版本B5.5及以上版本)	RBC	×10 ¹² /L	2.25 ±0.18	4.73 ±0.24	5.48 ±0.50
	HGB	g/L	60 ±4	135 ±6	170 ±8
	HCT	%	20.4 ±2.0	43.9 ±3.0	54.1 ±4.0
	MCV	fL	90.7 ±5.0	92.9 ±5.0	98.7 ±6.0
	MCH	pg	27.2 ±2.5	28.5 ±2.5	30.9 ±2.5
	MCHC	g/L	329 ±30	311 ±30	316 ±30
	RDW-CV	%	18.6 ±3.0	16.4 ±3.0	16.1 ±3.0
	RDW-SD	fL	54.8 ±10.0	52.5 ±10.0	54.5 ±12.0
	PLT	×10 ⁹ /L	59 ±20	251 ±40	506 ±60
	MPV	fL	8.2 ±3.0	8.5 ±3.0	8.3 ±3.0
	PDW	fL	8.8 ±3.0	8.9 ±3.0	9.0 ±3.0
	PCT	%	0.048 ±0.048	0.213 ±0.100	0.420 ±0.200
	P-LCR	%	17.1 ±8.0	17.9 ±8.0	16.6 ±8.0
	P-LCC	×10 ⁹ /L	14 ±14	45 ±25	84 ±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND DF50 DF51 DF52 DF53 DF55 DF56 (适用于技术文件 版本A11.2至A11.6 版本)	WBC	×10 ⁹ /L	3.72 ±0.50	8.87 ±1.00	19.58 ±2.50
	Neu%	%	55.1 ±9.0	60.0 ±8.0	65.0 ±7.0
	Lym%	%	35.0 ±9.0	28.7 ±8.0	19.5 ±6.0
	Mon%	%	4.6 ±4.0	3.8 ±3.8	5.4 ±5.4
	Eos%	%	5.3 ±5.0	7.5 ±6.0	10.1 ±7.0
	Bas%	%	6.8 ±6.8	6.5 ±6.5	3.3 ±3.3
	Neu#	×10 ⁹ /L	2.05 ±0.40	5.31 ±0.70	12.72 ±1.40
	Lym#	×10 ⁹ /L	1.30 ±0.40	2.55 ±0.70	3.82 ±1.10
	Mon#	×10 ⁹ /L	0.17 ±0.14	0.34 ±0.34	1.06 ±1.06
	Eos#	×10 ⁹ /L	0.20 ±0.15	0.67 ±0.50	1.98 ±1.30
	Bas#	×10 ⁹ /L	0.25 ±0.25	0.58 ±0.58	0.65 ±0.65
	RBC	×10 ¹² /L	2.31 ±0.18	4.53 ±0.24	5.25 ±0.50
	HGB	g/L	59 ±4	133 ±6	168 ±8
	HCT	%	18.9 ±2.0	40.7 ±3.0	50.2 ±4.0
	MCV	fL	82.0 ±5.0	89.9 ±5.0	95.7 ±6.0
	MCH	pg	25.4 ±2.5	29.2 ±2.5	31.7 ±2.5
	MCHC	g/L	310 ±30	325 ±30	333 ±30
	RDW-CV	%	15.4 ±3.0	14.0 ±3.0	13.7 ±3.0
	RDW-SD	fL	53.3 ±10.0	52.1 ±10.0	54.1 ±12.0
	PLT	×10 ⁹ /L	62 ±20	247 ±40	481 ±60
	MPV	fL	8.4 ±3.0	8.4 ±3.0	8.5 ±3.0
	PDW	fL	8.8 ±3.0	10.6 ±3.0	10.8 ±3.0
	PCT	%	0.052 ±0.050	0.207 ±0.100	0.409 ±0.200
	P-LCR	%	22.8 ±8.0	25.3 ±8.0	25.4 ±8.0
	P-LCC	×10 ⁹ /L	15 ±15	62 ±25	123 ±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND DF50 DF51 DF52 DF53 DF55 DF56 (适用于技术文件 版本A12.0及以上 版本)	WBC	×10 ⁹ /L	3.91 ±0.50	9.30 ±1.00	20.53 ±2.50
	Neu%	%	52.5 ±9.0	57.0 ±8.0	62.7 ±7.0
	Lym%	%	37.4 ±9.0	30.3 ±8.0	20.0 ±6.0
	Mon%	%	5.0 ±4.0	5.4 ±5.0	7.3 ±6.0
	Eos%	%	5.1 ±5.0	7.3 ±6.0	10.0 ±7.0
	Bas%	%	5.1 ±5.1	4.5 ±4.5	2.4 ±2.4
	Neu#	×10 ⁹ /L	2.05 ±0.40	5.30 ±0.70	12.87 ±1.40
	Lym#	×10 ⁹ /L	1.46 ±0.40	2.82 ±0.70	4.11 ±1.10
	Mon#	×10 ⁹ /L	0.20 ±0.14	0.50 ±0.50	1.50 ±1.10
	Eos#	×10 ⁹ /L	0.20 ±0.15	0.68 ±0.50	2.05 ±1.30
	Bas#	×10 ⁹ /L	0.20 ±0.20	0.42 ±0.42	0.49 ±0.49
	RBC	×10 ¹² /L	2.34 ±0.18	4.60 ±0.24	5.30 ±0.50
	HGB	g/L	61 ±4	136 ±6	171 ±8
	HCT	%	19.0 ±2.0	40.0 ±3.0	48.8 ±4.0
	MCV	fL	81.0 ±5.0	87.0 ±5.0	92.0 ±6.0
	MCH	pg	25.6 ±2.5	29.3 ±2.5	31.9 ±2.5
	MCHC	g/L	318 ±30	338 ±30	349 ±30
	RDW-CV	%	16.0 ±3.0	14.6 ±3.0	14.3 ±3.0
	RDW-SD	fL	55.0 ±10.0	54.3 ±10.0	56.5 ±12.0
	PLT	×10 ⁹ /L	58 ±20	240 ±40	466 ±60
	MPV	fL	7.9 ±3.0	7.9 ±3.0	8.1 ±3.0
	PDW	fL	8.1 ±3.0	9.8 ±3.0	10.2 ±3.0
	PCT	%	0.046 ±0.046	0.190 ±0.100	0.377 ±0.200
	P-LCR	%	19.1 ±8.0	22.2 ±8.0	23.0 ±8.0
	P-LCC	×10 ⁹ /L	15 ±15	53 ±25	107 ±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND DF50 DF51 DF52 DF53 DF55 DF56 (适用于技术文件 版本B1.0及以上版 本)	WBC	×10 ⁹ /L	3.80 ±0.50	9.01 ±1.00	19.83 ±2.50
	Neu%	%	52.5 ±9.0	58.8 ±8.0	64.5 ±7.0
	Lym%	%	34.9 ±9.0	28.6 ±8.0	18.9 ±6.0
	Mon%	%	8.0 ±4.0	5.4 ±5.0	6.8 ±6.0
	Eos%	%	4.6 ±4.6	7.2 ±6.0	9.8 ±7.0
	Bas%	%	5.9 ±5.9	5.0 ±5.0	2.8 ±2.8
	Neu#	×10 ⁹ /L	2.00 ±0.40	5.29 ±0.70	12.79 ±1.40
	Lym#	×10 ⁹ /L	1.33 ±0.40	2.58 ±0.70	3.75 ±1.10
	Mon#	×10 ⁹ /L	0.30 ±0.14	0.49 ±0.49	1.35 ±1.10
	Eos#	×10 ⁹ /L	0.17 ±0.15	0.65 ±0.50	1.94 ±1.30
	Bas#	×10 ⁹ /L	0.22 ±0.22	0.45 ±0.45	0.56 ±0.56
	RBC	×10 ¹² /L	2.30 ±0.18	4.52 ±0.24	5.22 ±0.50
	HGB	g/L	60 ±4	135 ±6	170 ±8
	HCT	%	18.8 ±2.0	40.0 ±3.0	48.8 ±4.0
	MCV	fL	81.8 ±5.0	88.5 ±5.0	93.4 ±6.0
	MCH	pg	26.1 ±2.5	29.9 ±2.5	32.6 ±2.5
	MCHC	g/L	321 ±30	341 ±30	351 ±30
	RDW-CV	%	15.8 ±3.0	14.3 ±3.0	14.0 ±3.0
	RDW-SD	fL	54.4 ±10.0	53.4 ±10.0	55.2 ±12.0
	PLT	×10 ⁹ /L	59 ±20	244 ±40	467 ±60
	MPV	fL	8.0 ±3.0	7.7 ±3.0	7.9 ±3.0
	PDW	fL	8.3 ±3.0	10.3 ±3.0	10.5 ±3.0
	PCT	%	0.047 ±0.047	0.188 ±0.100	0.369 ±0.200
	P-LCR	%	18.5 ±8.0	20.8 ±8.0	21.2 ±8.0
	P-LCC	×10 ⁹ /L	15 ±15	51 ±25	99 ±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND DH71 DH73 DH76 (适用于技术文件 版本A4.1至A10.3 版本) DH51 DH53 DH56 (适用于技术文件 版本A8.1及以上版 本)	WBC	×10 ⁹ /L	3.65 ±0.50	8.68 ±1.00	19.69 ±2.50
	Neu%	%	52.0 ±9.0	58.1 ±8.0	64.8 ±7.0
	Lym%	%	35.6 ±9.0	31.1 ±8.0	20.7 ±6.0
	Mon%	%	8.0 ±4.0	5.5 ±5.0	7.7 ±6.0
	Eos%	%	4.4 ±4.4	5.3 ±5.3	6.8 ±6.8
	Bas%	%	64.6 ±8.0	70.0 ±8.0	80.0 ±8.0
	Neu#	×10 ⁹ /L	1.90 ±0.40	5.04 ±0.70	12.75 ±1.40
	Lym#	×10 ⁹ /L	1.30 ±0.40	2.70 ±0.70	4.08 ±1.10
	Mon#	×10 ⁹ /L	0.29 ±0.14	0.48 ±0.48	1.52 ±1.10
	Eos#	×10 ⁹ /L	0.16 ±0.15	0.46 ±0.46	1.34 ±1.30
	Bas#	×10 ⁹ /L	2.36 ±0.30	6.08 ±0.70	15.75 ±1.50
	RBC	×10 ¹² /L	2.19 ±0.18	4.35 ±0.24	5.04 ±0.50
	HGB	g/L	59 ±4	135 ±6	168 ±8
	HCT	%	18.3 ±2.0	40.1 ±3.0	49.6 ±4.0
	MCV	fL	83.4 ±5.0	92.2 ±5.0	98.5 ±6.0
	MCH	pg	26.9 ±2.5	30.8 ±2.5	33.1 ±2.5
	MCHC	g/L	333 ±30	345 ±30	348 ±30
	RDW-CV	%	18.8 ±3.0	16.5 ±3.0	16.0 ±3.0
	RDW-SD	fL	53.9 ±10.0	51.8 ±10.0	53.3 ±12.0
	PLT	×10 ⁹ /L	46 ±20	237 ±40	498 ±60
	MPV	fL	9.2 ±3.0	8.6 ±3.0	8.6 ±3.0
	PDW	fL	8.0 ±3.0	9.4 ±3.0	9.6 ±3.0
	PCT	%	0.042 ±0.042	0.204 ±0.100	0.428 ±0.200
	P-LCR	%	18.0 ±8.0	18.4 ±8.0	17.5 ±8.0
	P-LCC	×10 ⁹ /L	14 ±14	44 ±25	87 ±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND DH71 DH73 DH76 (适用于技术文件 版本 A10.4版本)	WBC	×10 ⁹ /L	3.21 ±0.50	7.70 ±1.00	17.47 ±2.50
	Neu%	%	51.2 ±9.0	55.9 ±8.0	63.9 ±7.0
	Lym%	%	35.8 ±9.0	31.3 ±8.0	21.4 ±6.0
	Mon%	%	8.7 ±4.0	7.8 ±5.0	9.1 ±6.0
	Eos%	%	4.3 ±4.3	5.0 ±5.0	5.6 ±5.6
	Bas%	%	62.8 ±8.0	68.8 ±8.0	79.1 ±8.0
	Neu#	×10 ⁹ /L	1.64 ±0.40	4.30 ±0.70	11.16 ±1.40
	Lym#	×10 ⁹ /L	1.15 ±0.40	2.41 ±0.70	3.74 ±1.10
	Mon#	×10 ⁹ /L	0.28 ±0.14	0.60 ±0.50	1.59 ±1.10
	Eos#	×10 ⁹ /L	0.14 ±0.14	0.39 ±0.39	0.98 ±0.98
	Bas#	×10 ⁹ /L	2.02 ±0.30	5.30 ±0.70	13.82 ±1.50
	RBC	×10 ¹² /L	2.22 ±0.18	4.52 ±0.24	5.26 ±0.50
	HGB	g/L	57 ±4	133 ±6	165 ±8
	HCT	%	18.1 ±2.0	40.3 ±3.0	49.8 ±4.0
	MCV	fL	81.5 ±5.0	89.2 ±5.0	94.7 ±6.0
	MCH	pg	28.3 ±2.5	32.3 ±2.5	34.6 ±2.5
	MCHC	g/L	350 ±30	365 ±30	368 ±30
	RDW-CV	%	18.8 ±3.0	16.6 ±3.0	16.2 ±3.0
	RDW-SD	fL	55.8 ±10.0	53.7 ±10.0	55.5 ±12.0
	PLT	×10 ⁹ /L	45 ±20	236 ±40	502 ±60
	MPV	fL	9.1 ±3.0	8.6 ±3.0	8.5 ±3.0
	PDW	fL	7.9 ±3.0	9.2 ±3.0	9.3 ±3.0
	PCT	%	0.041 ±0.041	0.203 ±0.100	0.427 ±0.200
	P-LCR	%	17.7 ±8.0	18.2 ±8.0	17.2 ±8.0
	P-LCC	×10 ⁹ /L	14 ±14	43 ±25	87 ±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND DH71 DH73 DH76 (适用于技术文件 版本 A10.5及以上 版本)	WBC	×10 ⁹ /L	3.74 ±0.50	9.05 ±1.00	20.86 ±2.50
	Neu%	%	51.4 ±9.0	55.5 ±8.0	63.3 ±7.0
	Lym%	%	36.2 ±9.0	31.7 ±8.0	21.2 ±6.0
	Mon%	%	8.7 ±4.0	7.5 ±5.0	8.5 ±6.0
	Eos%	%	3.7 ±3.7	5.3 ±5.3	7.0 ±7.0
	Bas%	%	63.9 ±8.0	71.5 ±8.0	81.4 ±8.0
	Neu#	×10 ⁹ /L	1.92 ±0.40	5.02 ±0.70	13.21 ±1.40
	Lym#	×10 ⁹ /L	1.35 ±0.40	2.87 ±0.70	4.42 ±1.10
	Mon#	×10 ⁹ /L	0.33 ±0.14	0.68 ±0.50	1.77 ±1.10
	Eos#	×10 ⁹ /L	0.14 ±0.14	0.48 ±0.48	1.46 ±1.30
	Bas#	×10 ⁹ /L	2.39 ±0.30	6.47 ±0.70	16.98 ±1.50
	RBC	×10 ¹² /L	2.34 ±0.18	4.69 ±0.24	5.47 ±0.50
	HGB	g/L	59 ±4	135 ±6	168 ±8
	HCT	%	18.8 ±2.0	41.3 ±3.0	51.2 ±4.0
	MCV	fL	80.2 ±5.0	88.0 ±5.0	93.6 ±6.0
	MCH	pg	25.2 ±2.5	28.8 ±2.5	30.7 ±2.5
	MCHC	g/L	314 ±30	327 ±30	328 ±30
	RDW-CV	%	19.1 ±3.0	16.8 ±3.0	16.3 ±3.0
	RDW-SD	fL	50.5 ±10.0	48.4 ±10.0	49.7 ±12.0
	PLT	×10 ⁹ /L	44 ±20	242 ±40	513 ±60
	MPV	fL	8.4 ±3.0	7.6 ±3.0	7.6 ±3.0
	PDW	fL	6.2 ±3.0	7.4 ±3.0	7.3 ±3.0
	PCT	%	0.037 ±0.037	0.184 ±0.100	0.390 ±0.200
	P-LCR	%	13.8 ±8.0	13.1 ±8.0	12.6 ±8.0
	P-LCC	×10 ⁹ /L	12 ±12	32 ±25	65 ±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low		Normal		High	
DYMIND D2-CRP D7-CRP DH71CRP DH73CRP (适用于技术文件 版本A6.0至A6.5版 本)	WBC	×10 ⁹ /L	3.71	±0.50	8.83	±1.00	20.12	±2.50
	Neu%	%	52.2	±9.0	56.7	±8.0	63.8	±7.0
	Lym%	%	35.6	±9.0	31.1	±8.0	21.4	±6.0
	Mon%	%	8.5	±4.0	7.4	±5.0	8.2	±6.0
	Eos%	%	3.7	±3.7	4.8	±4.8	6.6	±6.6
	Bas%	%	65.3	±8.0	70.8	±8.0	80.8	±8.0
	Neu#	×10 ⁹ /L	1.93	±0.40	5.01	±0.70	12.83	±1.40
	Lym#	×10 ⁹ /L	1.32	±0.40	2.75	±0.70	4.31	±1.10
	Mon#	×10 ⁹ /L	0.32	±0.14	0.65	±0.50	1.65	±1.10
	Eos#	×10 ⁹ /L	0.14	±0.14	0.42	±0.42	1.33	±1.30
	Bas#	×10 ⁹ /L	2.42	±0.30	6.25	±0.70	16.26	±1.50
	RBC	×10 ¹² /L	2.34	±0.18	4.66	±0.24	5.39	±0.50
	HGB	g/L	60	±4	137	±6	170	±8
	HCT	%	19.2	±2.0	41.5	±3.0	50.9	±4.0
	MCV	fL	82.1	±5.0	89.0	±5.0	94.4	±6.0
	MCH	pg	25.8	±2.5	29.4	±2.5	31.5	±2.5
	MCHC	g/L	317	±30	334	±30	336	±30
	RDW-CV	%	19.1	±3.0	17.2	±3.0	16.4	±3.0
	RDW-SD	fL	56.2	±10.0	55.1	±10.0	56.2	±12.0
	PLT	×10 ⁹ /L	51	±20	250	±40	536	±60
	MPV	fL	8.9	±3.0	8.5	±3.0	8.5	±3.0
	PDW	fL	7.6	±3.0	9.2	±3.0	9.4	±3.0
	PCT	%	0.045	±0.045	0.213	±0.100	0.456	±0.200
	P-LCR	%	16.5	±8.0	17.8	±8.0	17.1	±8.0
	P-LCC	×10 ⁹ /L	14	±14	45	±25	92	±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND D2-CRP D7-CRP DH71CRP DH73CRP (适用于技术文件 版本A6.6及以上版本)	WBC	×10 ⁹ /L	3.91 ±0.50	9.27 ±1.00	21.16 ±2.50
	Neu%	%	50.6 ±9.0	56.1 ±8.0	63.0 ±7.0
	Lym%	%	36.3 ±9.0	31.6 ±8.0	21.4 ±6.0
	Mon%	%	8.8 ±4.0	6.4 ±5.0	7.8 ±6.0
	Eos%	%	4.3 ±4.3	5.9 ±5.9	7.8 ±7.0
	Bas%	%	65.4 ±8.0	71.2 ±8.0	81.2 ±8.0
	Neu#	×10 ⁹ /L	1.98 ±0.40	5.20 ±0.70	13.33 ±1.40
	Lym#	×10 ⁹ /L	1.42 ±0.40	2.93 ±0.70	4.53 ±1.10
	Mon#	×10 ⁹ /L	0.34 ±0.14	0.59 ±0.50	1.65 ±1.10
	Eos#	×10 ⁹ /L	0.17 ±0.15	0.55 ±0.50	1.65 ±1.30
	Bas#	×10 ⁹ /L	2.56 ±0.30	6.60 ±0.70	17.18 ±1.50
	RBC	×10 ¹² /L	2.38 ±0.18	4.73 ±0.24	5.45 ±0.50
	HGB	g/L	60 ±4	135 ±6	167 ±8
	HCT	%	19.2 ±2.0	41.6 ±3.0	51.0 ±4.0
	MCV	fL	80.5 ±5.0	88.0 ±5.0	93.5 ±6.0
	MCH	pg	25.2 ±2.5	28.5 ±2.5	30.6 ±2.5
	MCHC	g/L	313 ±30	325 ±30	327 ±30
	RDW-CV	%	18.7 ±3.0	16.5 ±3.0	16.0 ±3.0
	RDW-SD	fL	55.2 ±10.0	52.8 ±10.0	54.5 ±12.0
	PLT	×10 ⁹ /L	50 ±20	241 ±40	508 ±60
	MPV	fL	8.5 ±3.0	7.9 ±3.0	7.9 ±3.0
	PDW	fL	6.6 ±3.0	7.8 ±3.0	7.8 ±3.0
	PCT	%	0.043 ±0.043	0.190 ±0.100	0.401 ±0.200
	P-LCR	%	14.3 ±8.0	14.6 ±8.0	14.2 ±8.0
	P-LCC	×10 ⁹ /L	11 ±11	35 ±25	72 ±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND D6-CRP DH76CRP (适用于技术文件 版本B1.0至B2.0版 本)	WBC	×10 ⁹ /L	3.63 ±0.50	8.58 ±1.00	19.55 ±2.50
	Neu%	%	51.4 ±9.0	56.3 ±8.0	64.1 ±7.0
	Lym%	%	35.5 ±9.0	31.1 ±8.0	21.0 ±6.0
	Mon%	%	9.8 ±4.0	7.9 ±5.0	8.4 ±6.0
	Eos%	%	3.3 ±3.3	4.7 ±4.7	6.5 ±6.5
	Bas%	%	65.0 ±8.0	70.8 ±8.0	80.7 ±8.0
	Neu#	×10 ⁹ /L	1.86 ±0.40	4.83 ±0.70	12.53 ±1.40
	Lym#	×10 ⁹ /L	1.29 ±0.40	2.67 ±0.70	4.11 ±1.10
	Mon#	×10 ⁹ /L	0.36 ±0.14	0.68 ±0.50	1.64 ±1.10
	Eos#	×10 ⁹ /L	0.12 ±0.12	0.40 ±0.40	1.27 ±1.27
	Bas#	×10 ⁹ /L	2.36 ±0.30	6.07 ±0.70	15.78 ±1.50
	RBC	×10 ¹² /L	2.34 ±0.18	4.63 ±0.24	5.35 ±0.50
	HGB	g/L	62 ±4	136 ±6	167 ±8
	HCT	%	19.0 ±2.0	40.8 ±3.0	50.1 ±4.0
	MCV	fL	81.0 ±5.0	88.1 ±5.0	93.6 ±6.0
	MCH	pg	26.7 ±2.5	29.4 ±2.5	31.4 ±2.5
	MCHC	g/L	332 ±30	336 ±30	338 ±30
	RDW-CV	%	18.9 ±3.0	16.7 ±3.0	16.2 ±3.0
	RDW-SD	fL	55.2 ±10.0	53.2 ±10.0	54.9 ±12.0
	PLT	×10 ⁹ /L	56 ±20	240 ±40	513 ±60
	MPV	fL	7.6 ±3.0	7.9 ±3.0	8.0 ±3.0
	PDW	fL	7.8 ±3.0	8.2 ±3.0	8.3 ±3.0
	PCT	%	0.043 ±0.043	0.190 ±0.100	0.410 ±0.200
	P-LCR	%	14.2 ±8.0	14.6 ±8.0	14.4 ±8.0
	P-LCC	×10 ⁹ /L	14 ±14	35 ±25	74 ±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND D6-CRP DH76CRP (适用于技术文件 版本B2.2及以上版本)	WBC	×10 ⁹ /L	3.88 ±0.50	9.14 ±1.00	20.69 ±2.50
	Neu%	%	51.6 ±9.0	55.8 ±8.0	63.1 ±7.0
	Lym%	%	35.2 ±9.0	30.5 ±8.0	20.7 ±6.0
	Mon%	%	8.5 ±4.0	7.4 ±5.0	8.2 ±6.0
	Eos%	%	4.7 ±4.7	6.3 ±6.0	8.0 ±7.0
	Bas%	%	65.0 ±8.0	70.9 ±8.0	81.0 ±8.0
	Neu#	×10 ⁹ /L	2.00 ±0.40	5.09 ±0.70	13.05 ±1.40
	Lym#	×10 ⁹ /L	1.37 ±0.40	2.79 ±0.70	4.28 ±1.10
	Mon#	×10 ⁹ /L	0.33 ±0.14	0.68 ±0.50	1.70 ±1.10
	Eos#	×10 ⁹ /L	0.18 ±0.15	0.58 ±0.50	1.66 ±1.30
	Bas#	×10 ⁹ /L	2.52 ±0.30	6.48 ±0.70	16.76 ±1.50
	RBC	×10 ¹² /L	2.41 ±0.18	4.78 ±0.24	5.54 ±0.50
	HGB	g/L	61 ±4	135 ±6	167 ±8
	HCT	%	19.8 ±2.0	42.7 ±3.0	52.6 ±4.0
	MCV	fL	82.2 ±5.0	89.4 ±5.0	94.9 ±6.0
	MCH	pg	26.2 ±2.5	28.9 ±2.5	30.9 ±2.5
	MCHC	g/L	320 ±30	325 ±30	328 ±30
	RDW-CV	%	18.8 ±3.0	16.5 ±3.0	16.0 ±3.0
	RDW-SD	fL	56.5 ±10.0	54.2 ±10.0	55.7 ±12.0
	PLT	×10 ⁹ /L	53 ±20	253 ±40	539 ±60
	MPV	fL	7.8 ±3.0	8.3 ±3.0	8.1 ±3.0
	PDW	fL	8.1 ±3.0	8.5 ±3.0	8.7 ±3.0
	PCT	%	0.041 ±0.041	0.210 ±0.100	0.437 ±0.200
	P-LCR	%	15.0 ±8.0	16.6 ±8.0	15.2 ±8.0
	P-LCC	×10 ⁹ /L	14 ±14	42 ±25	82 ±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low		Normal		High	
DYMIND DF50 Vet DF52 Vet DF55 Vet DF56 Vet (适用于技术文件 版本A8.0及以上版 本、B1.0及以上版 本)	WBC	×10 ⁹ /L	3.74	±0.50	8.87	±1.00	19.80	±2.50
	Neu%	%	53.7	±9.0	59.0	±8.0	65.3	±7.0
	Lym%	%	37.0	±9.0	29.7	±8.0	15.8	±10.0
	Mon%	%	3.5	±3.5	3.4	±3.4	8.7	±8.0
	Eos%	%	5.8	±5.0	7.9	±6.0	10.2	±7.0
	Bas%	%	2.9	±2.9	2.9	±2.9	1.7	±1.7
	Neu#	×10 ⁹ /L	2.01	±0.40	5.24	±0.70	12.93	±1.40
	Lym#	×10 ⁹ /L	1.38	±0.40	2.63	±0.70	3.13	±2.50
	Mon#	×10 ⁹ /L	0.13	±0.13	0.30	±0.30	1.72	±1.50
	Eos#	×10 ⁹ /L	0.22	±0.15	0.70	±0.50	2.02	±1.30
	Bas#	×10 ⁹ /L	0.11	±0.11	0.26	±0.26	0.34	±0.34
	RBC	×10 ¹² /L	2.33	±0.18	4.58	±0.24	5.26	±0.50
	HGB	g/L	59	±4	135	±6	170	±8
	HCT	%	19.0	±2.0	41.1	±3.0	50.4	±4.0
	MCV	fL	81.4	±5.0	89.7	±5.0	95.8	±6.0
	MCH	pg	25.3	±2.5	29.2	±2.5	32.0	±2.5
	MCHC	g/L	316	±30	332	±30	341	±30
	RDW-CV	%	14.7	±3.0	13.4	±3.0	13.1	±3.0
	RDW-SD	fL	50.4	±10.0	49.6	±10.0	51.4	±12.0
	PLT	×10 ⁹ /L	59	±20	239	±40	459	±60
	MPV	fL	7.8	±3.0	7.6	±3.0	7.8	±3.0
	PDW	fL	7.7	±3.0	9.7	±3.0	9.8	±3.0
	PCT	%	0.046	±0.046	0.182	±0.100	0.358	±0.200

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND DF1-CRP DF3-CRP DF5-CRP DF50CRP DF52CRP DF53CRP (适用于技术文件 版本A6.1及以上版 本)	WBC	×10 ⁹ /L	4.05 ±0.50	9.74 ±1.00	22.13 ±2.50
	Neu%	%	51.5 ±9.0	56.4 ±8.0	62.6 ±7.0
	Lym%	%	35.6 ±9.0	29.9 ±8.0	19.8 ±6.0
	Mon%	%	8.0 ±4.0	6.4 ±5.0	7.7 ±6.0
	Eos%	%	4.9 ±4.9	7.3 ±6.0	9.9 ±7.0
	Bas%	%	5.1 ±5.1	4.5 ±4.5	2.4 ±2.4
	Neu#	×10 ⁹ /L	2.09 ±0.40	5.50 ±0.70	13.86 ±1.40
	Lym#	×10 ⁹ /L	1.44 ±0.40	2.91 ±0.70	4.38 ±1.10
	Mon#	×10 ⁹ /L	0.32 ±0.14	0.62 ±0.50	1.70 ±1.10
	Eos#	×10 ⁹ /L	0.20 ±0.15	0.71 ±0.50	2.19 ±1.30
	Bas#	×10 ⁹ /L	0.21 ±0.21	0.44 ±0.44	0.53 ±0.53
	RBC	×10 ¹² /L	2.29 ±0.18	4.48 ±0.24	5.18 ±0.50
	HGB	g/L	61 ±4	136 ±6	170 ±8
	HCT	%	18.4 ±2.0	38.4 ±3.0	47.2 ±4.0
	MCV	fL	80.2 ±5.0	85.7 ±5.0	91.1 ±6.0
	MCH	pg	26.6 ±2.5	30.4 ±2.5	32.8 ±2.5
	MCHC	g/L	332 ±30	354 ±30	360 ±30
	RDW-CV	%	15.6 ±3.0	14.4 ±3.0	13.1 ±3.0
	RDW-SD	fL	54.9 ±10.0	54.3 ±10.0	56.3 ±12.0
	PLT	×10 ⁹ /L	50 ±20	255 ±40	532 ±60
	MPV	fL	7.9 ±3.0	7.4 ±3.0	7.5 ±3.0
	PDW	fL	6.9 ±3.0	9.0 ±3.0	9.1 ±3.0
	PCT	%	0.040 ±0.040	0.189 ±0.100	0.399 ±0.200
	P-LCR	%	16.5 ±8.0	18.7 ±8.0	19.1 ±8.0
	P-LCC	×10 ⁹ /L	13 ±13	48 ±25	102 ±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND DM71X DM72X DM78X DM79X (适用于技术文件 版本A1.0及以上版本)	WBC	×10 ⁹ /L	3.62 ±0.50	8.64 ±1.00	19.61 ±2.50
	Neu%	%	51.1 ±9.0	55.8 ±8.0	63.3 ±7.0
	Lym%	%	34.9 ±9.0	30.1 ±8.0	20.3 ±6.0
	Mon%	%	9.6 ±4.0	8.2 ±5.0	8.5 ±6.0
	Eos%	%	4.4 ±4.4	5.9 ±5.9	7.9 ±7.0
	Bas%	%	64.2 ±8.0	70.0 ±8.0	80.6 ±8.0
	Neu#	×10 ⁹ /L	1.85 ±0.40	4.82 ±0.70	12.41 ±1.40
	Lym#	×10 ⁹ /L	1.26 ±0.40	2.60 ±0.70	3.98 ±1.10
	Mon#	×10 ⁹ /L	0.35 ±0.14	0.71 ±0.50	1.67 ±1.10
	Eos#	×10 ⁹ /L	0.16 ±0.15	0.51 ±0.50	1.55 ±1.30
	Bas#	×10 ⁹ /L	2.32 ±0.30	6.05 ±0.70	15.81 ±1.50
	RBC	×10 ¹² /L	2.36 ±0.18	4.68 ±0.24	5.44 ±0.50
	HGB	g/L	62 ±4	136 ±6	168 ±8
	HCT	%	18.8 ±2.0	41.0 ±3.0	51.0 ±4.0
	MCV	fL	79.5 ±5.0	87.7 ±5.0	93.7 ±6.0
	MCH	pg	26.8 ±2.5	29.5 ±2.5	31.3 ±2.5
	MCHC	g/L	335 ±30	336 ±30	334 ±30
	RDW-CV	%	19.0 ±3.0	17.2 ±3.0	16.5 ±3.0
	RDW-SD	fL	55.5 ±10.0	54.3 ±10.0	55.8 ±12.0
	PLT	×10 ⁹ /L	51 ±20	245 ±40	524 ±60
	MPV	fL	7.7 ±3.0	8.0 ±3.0	7.9 ±3.0
	PDW	fL	8.0 ±3.0	8.4 ±3.0	8.3 ±3.0
	PCT	%	0.039 ±0.039	0.196 ±0.100	0.414 ±0.200
	P-LCR	%	13.8 ±8.0	14.9 ±8.0	14.1 ±8.0
	P-LCC	×10 ⁹ /L	13 ±13	36 ±25	74 ±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low	Normal	High
DYMIND UN71 Vet UN73 Vet (适用于技术文件 版本A1.0及以上版本)	WBC	×10 ⁹ /L	3.51 ±0.50	8.55 ±1.00	19.86 ±2.50
	Neu%	%	51.7 ±9.0	56.5 ±8.0	65.0 ±7.0
	Lym%	%	35.2 ±9.0	30.4 ±8.0	20.9 ±6.0
	Mon%	%	9.4 ±4.0	8.2 ±5.0	8.7 ±6.0
	Eos%	%	3.7 ±3.7	4.9 ±4.9	5.4 ±5.4
	Bas%	%	0.9 ±0.9	0.9 ±0.9	0.9 ±0.9
	Neu#	×10 ⁹ /L	1.81 ±0.40	4.83 ±0.70	12.91 ±1.40
	Lym#	×10 ⁹ /L	1.24 ±0.40	2.60 ±0.70	4.15 ±1.10
	Mon#	×10 ⁹ /L	0.33 ±0.14	0.70 ±0.50	1.73 ±1.10
	Eos#	×10 ⁹ /L	0.13 ±0.13	0.42 ±0.42	1.07 ±1.07
	Bas#	×10 ⁹ /L	0.03 ±0.03	0.08 ±0.08	0.18 ±0.18
	RBC	×10 ¹² /L	2.34 ±0.18	4.61 ±0.24	5.29 ±0.50
	HGB	g/L	62 ±4	135 ±6	167 ±8
	HCT	%	18.9 ±2.0	41.1 ±3.0	50.4 ±4.0
	MCV	fL	80.9 ±5.0	89.2 ±5.0	95.3 ±6.0
	MCH	pg	26.8 ±2.5	29.7 ±2.5	32.1 ±2.5
	MCHC	g/L	336 ±30	337 ±30	342 ±30
	RDW-CV	%	19.1 ±3.0	16.9 ±3.0	16.3 ±3.0
	RDW-SD	fL	48.7 ±10.0	47.0 ±10.0	48.2 ±12.0
	PLT	×10 ⁹ /L	42 ±20	233 ±40	477 ±60
	MPV	fL	8.8 ±3.0	7.6 ±3.0	7.7 ±3.0
	PDW	fL	6.0 ±3.0	7.5 ±3.0	7.5 ±3.0
	PCT	%	0.037 ±0.037	0.177 ±0.100	0.367 ±0.200
	P-LCR	%	13.7 ±8.0	13.5 ±8.0	13.3 ±8.0
	P-LCC	×10 ⁹ /L	12 ±12	31 ±25	63 ±35

HC5D-B

(Cat No.1510)

For MINDRAY
HEMATOLOGY CONTROLS

CONTROL

LOT E5221

EXP. 2025-09-16

适用机型	参数	单位	Low		Normal		High	
DYMIND DM60 Vet DM61 Vet DM62 Vet DM63 Vet (适用于技术文件 版本A1.4及以上版本)	WBC	×10 ⁹ /L	3.45	±0.50	8.22	±1.00	18.16	±2.50
	Neu%	%	53.1	±9.0	56.3	±8.0	62.8	±7.0
	Lym%	%	34.7	±9.0	29.0	±8.0	19.0	±6.0
	Mon%	%	7.2	±4.0	7.2	±5.0	8.2	±6.0
	Eos%	%	5.0	±5.0	7.5	±6.0	10.0	±7.0
	Bas%	%	5.5	±5.5	4.5	±4.5	2.3	±2.3
	Neu#	×10 ⁹ /L	1.83	±0.40	4.63	±0.70	11.40	±1.40
	Lym#	×10 ⁹ /L	1.20	±0.40	2.38	±0.70	3.45	±1.10
	Mon#	×10 ⁹ /L	0.25	±0.14	0.59	±0.50	1.49	±1.10
	Eos#	×10 ⁹ /L	0.17	±0.15	0.62	±0.50	1.82	±1.30
	Bas#	×10 ⁹ /L	0.19	±0.19	0.37	±0.37	0.42	±0.42
	RBC	×10 ¹² /L	2.17	±0.18	4.21	±0.24	4.87	±0.50
	HGB	g/L	58	±4	133	±6	168	±8
	HCT	%	17.8	±2.0	37.4	±3.0	45.8	±4.0
	MCV	fL	81.8	±5.0	88.8	±5.0	94.0	±6.0
	MCH	pg	26.7	±2.5	30.8	±2.5	33.9	±2.5
	MCHC	g/L	334	±30	357	±30	370	±30
	RDW-CV	%	14.6	±3.0	13.3	±3.0	13.0	±3.0
	RDW-SD	fL	53.1	±10.0	52.4	±10.0	54.4	±12.0
	PLT	×10 ⁹ /L	54	±20	236	±40	455	±60
	MPV	fL	7.9	±3.0	7.6	±3.0	7.7	±3.0
	PDW	fL	7.2	±3.0	9.4	±3.0	9.7	±3.0
	PCT	%	0.043	±0.043	0.179	±0.100	0.350	±0.200